

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092022\
 Data File : VX031505.D
 Acq On : 20 Sep 2022 18:53
 Operator : JC/MD
 Sample : N4641-12
 Misc : 6.02G/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 D22021-6

Quant Time: Sep 21 04:55:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 20 09:45:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	69013	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	112261	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216606	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	178908	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72672	57.473	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 114.940%	
35) Dibromofluoromethane	5.385	113	55439	47.746	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.500%	
50) Toluene-d8	8.647	98	328220	73.385	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 146.760%#	
62) 4-Bromofluorobenzene	11.085	95	180315	119.020	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 238.040%#	
Target Compounds						
					Qvalue	
15) Acrylonitrile	3.081	53	7894	11.852	ug/l	# 42
31) Cyclohexane	5.465	56	228632	121.794	ug/l	# 89
39) Methylcyclohexane	7.379	83	546672	248.988	ug/l	97
40) Benzene	6.031	78	40658	7.637	ug/l	98
52) Toluene	8.720	92	1947	0.561	ug/l	98
67) Ethyl Benzene	10.195	91	639613	54.233	ug/l	95
68) m/p-Xylenes	10.305	106	132788	28.017	ug/l	96
69) o-Xylene	10.647	106	9668	2.137	ug/l	84
73) Isopropylbenzene	10.964	105	376432	28.176	ug/l	97
78) n-propylbenzene	11.305	91	1665159	108.334	ug/l	93
80) 1,3,5-Trimethylbenzene	11.457	105	3116098	274.834	ug/l	93
84) 1,2,4-Trimethylbenzene	11.756	105	2868663	256.906	ug/l	93
85) sec-Butylbenzene	11.890	105	265076	19.060	ug/l	95
86) p-Isopropyltoluene	12.012	119	135507	11.539	ug/l	98
89) n-Butylbenzene	12.329	91	1349995	134.954	ug/l	# 16
95) Naphthalene	13.780	128	2824569	220.229	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092022\
Data File : VX031505.D
Acq On : 20 Sep 2022 18:53
Operator : JC/MD
Sample : N4641-12
Misc : 6.02G/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
D22021-6

Quant Time: Sep 21 04:55:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 20 09:45:29 2022
Response via : Initial Calibration

