



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092022\
 Data File : VX031484.D
 Acq On : 20 Sep 2022 09:31
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091922W.M
 Title : SW846 8260
 Last Update : Tue Sep 20 09:45:29 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.1	4588	PASS
75	95	30	60	58.6	11141	PASS
95	95	100	100	100.0	19003	PASS
96	95	5	9	7.6	1451	PASS
173	174	0.00	2	1.2	204	PASS
174	95	50	100	88.2	16760	PASS
175	174	5	9	8.5	1430	PASS
176	174	95	101	99.0	16595	PASS
177	176	5	9	6.7	1112	PASS