

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031651.D
 Acq On : 26 Sep 2022 16:17
 Operator : JC/MD
 Sample : VX0926WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0926WBS01

Quant Time: Sep 27 01:53:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	53428	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	79195	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	95021	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78995	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	54692	48.231	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.460%
35) Dibromofluoromethane	5.379	113	47158	49.975	ug/l	-0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.960%
50) Toluene-d8	8.647	98	179342	52.423	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	104.840%
62) 4-Bromofluorobenzene	11.079	95	68724	54.810	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.620%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.167	85	14237	21.895	ug/l	98
3) Chloromethane	1.288	50	14396	23.408	ug/l	94
4) Vinyl Chloride	1.368	62	19457	21.743	ug/l	100
5) Bromomethane	1.612	94	24164	27.483	ug/l	98
6) Chloroethane	1.685	64	23881	21.771	ug/l	100
7) Trichlorofluoromethane	1.886	101	49309	20.341	ug/l	97
8) Diethyl Ether	2.130	74	16078	23.624	ug/l	80
9) 1,1,2-Trichlorotrifluo...	2.331	101	14735	20.748	ug/l	93
10) Methyl Iodide	2.447	142	12553	18.032	ug/l	99
11) Tert butyl alcohol	2.995	59	15726	89.335	ug/l	99
12) 1,1-Dichloroethene	2.313	96	13071	22.428	ug/l	89
13) Acrolein	2.239	56	8224	104.842	ug/l	97
14) Allyl chloride	2.660	41	18375	19.665	ug/l	92
15) Acrylonitrile	3.069	53	39179	99.554	ug/l	99
16) Acetone	2.392	43	32401	112.985	ug/l	92
17) Carbon Disulfide	2.508	76	25982	20.938	ug/l	98
18) Methyl Acetate	2.703	43	20861	20.472	ug/l	94
19) Methyl tert-butyl Ether	3.111	73	49216	19.514	ug/l	96
20) Methylene Chloride	2.788	84	16208	21.643	ug/l	85
21) trans-1,2-Dichloroethene	3.087	96	15158	19.306	ug/l	85
22) Diisopropyl ether	3.758	45	47403	19.711	ug/l	87
23) Vinyl Acetate	3.721	43	194432	102.230	ug/l	96
24) 1,1-Dichloroethane	3.605	63	26991	19.468	ug/l	95
25) 2-Butanone	4.562	43	54536	100.756	ug/l	94
26) 2,2-Dichloropropane	4.477	77	20561	18.272	ug/l	95
27) cis-1,2-Dichloroethene	4.489	96	18953	19.711	ug/l	88
28) Bromochloromethane	4.898	49	13368	20.611	ug/l	86
29) Tetrahydrofuran	5.007	42	33587	98.778	ug/l	88
30) Chloroform	5.093	83	32871	19.272	ug/l	95
31) Cyclohexane	5.477	56	21857	20.272	ug/l	89
32) 1,1,1-Trichloroethane	5.379	97	27365	19.372	ug/l	99
36) 1,1-Dichloropropene	5.690	75	21764	20.083	ug/l	99
37) Ethyl Acetate	4.709	43	22450	20.538	ug/l	99
38) Carbon Tetrachloride	5.672	117	24272	19.874	ug/l	94
39) Methylcyclohexane	7.379	83	23923	19.375	ug/l	90
40) Benzene	6.038	78	65714	19.744	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031651.D
 Acq On : 26 Sep 2022 16:17
 Operator : JC/MD
 Sample : VX0926WBS01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0926WBS01

Quant Time: Sep 27 01:53:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	11219	20.171	ug/l	91
42) 1,2-Dichloroethane	6.086	62	26012	20.208	ug/l	98
43) Isopropyl Acetate	6.336	43	35212	20.468	ug/l	99
44) Trichloroethene	7.123	130	18951	20.159	ug/l	88
45) 1,2-Dichloropropane	7.428	63	16689	20.661	ug/l	98
46) Dibromomethane	7.580	93	13772	21.230	ug/l	94
47) Bromodichloromethane	7.818	83	24527	20.952	ug/l	96
48) Methyl methacrylate	7.696	41	16450	21.742	ug/l #	86
49) 1,4-Dioxane	7.690	88	8673	448.324	ug/l	92
51) 4-Methyl-2-Pentanone	8.574	43	119518	109.756	ug/l	97
52) Toluene	8.720	92	47370	21.351	ug/l	98
53) t-1,3-Dichloropropene	8.976	75	23884	18.884	ug/l	94
54) cis-1,3-Dichloropropene	8.366	75	24713	20.314	ug/l	100
55) 1,1,2-Trichloroethane	9.153	97	19398	20.308	ug/l	96
56) Ethyl methacrylate	9.116	69	27090	22.166	ug/l	94
57) 1,3-Dichloropropane	9.311	76	30892	20.421	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.238	63	63835	106.075	ug/l	91
59) 2-Hexanone	9.433	43	94903	116.808	ug/l	100
60) Dibromochloromethane	9.519	129	19403	18.401	ug/l	96
61) 1,2-Dibromoethane	9.610	107	21139	21.801	ug/l	100
64) Tetrachloroethene	9.275	164	19855	19.549	ug/l	94
65) Chlorobenzene	10.080	112	54356	20.533	ug/l	94
66) 1,1,1,2-Tetrachloroethane	10.165	131	19490	20.462	ug/l	98
67) Ethyl Benzene	10.195	91	92036	20.438	ug/l	99
68) m/p-Xylenes	10.299	106	76960	42.534	ug/l	99
69) o-Xylene	10.640	106	37140	21.000	ug/l	95
70) Styrene	10.659	104	62841	21.829	ug/l	97
71) Bromoform	10.799	173	14058	18.632	ug/l #	98
73) Isopropylbenzene	10.964	105	97464	19.078	ug/l	100
74) N-amyl acetate	10.842	43	33521	20.228	ug/l	95
75) 1,1,2,2-Tetrachloroethane	11.213	83	31974	19.698	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	32309	22.561	ug/l	88
77) Bromobenzene	11.201	156	24575	18.923	ug/l	95
78) n-propylbenzene	11.305	91	116326	20.463	ug/l	100
79) 2-Chlorotoluene	11.366	91	70198	19.997	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	88489	20.434	ug/l	100
81) trans-1,4-Dichloro-2-b...	11.018	75	6891	18.423	ug/l #	79
82) 4-Chlorotoluene	11.457	91	84732	20.893	ug/l	99
83) tert-Butylbenzene	11.713	119	82928	19.858	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	89314	20.836	ug/l	100
85) sec-Butylbenzene	11.890	105	109771	20.683	ug/l	100
86) p-Isopropyltoluene	12.012	119	94332	20.611	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	50433	20.032	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	53115	20.155	ug/l	98
89) n-Butylbenzene	12.335	91	78569	21.015	ug/l	99
90) Hexachloroethane	12.543	117	14628	19.932	ug/l	95
91) 1,2-Dichlorobenzene	12.335	146	51858	20.568	ug/l	98
92) 1,2-Dibromo-3-Chloropr...	12.945	75	6336	19.479	ug/l	92
93) 1,2,4-Trichlorobenzene	13.591	180	28567	19.375	ug/l	98
94) Hexachlorobutadiene	13.725	225	11099	19.305	ug/l	99
95) Naphthalene	13.774	128	98530	20.006	ug/l	100
96) 1,2,3-Trichlorobenzene	13.963	180	29275	19.346	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
Data File : VX031651.D
Acq On : 26 Sep 2022 16:17
Operator : JC/MD
Sample : VX0926WBS01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Quant Time: Sep 27 01:53:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Sep 27 01:53:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01

Tgt Ion:168 Resp: 53428
Ion Ratio Lower Upper
168 100
99 56.2 48.8 73.2






#2
Dichlorodifluoromethane
Concen: 21.895 ug/l
RT: 1.167 min Scan# 13
Delta R.T. 0.001 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 85 Resp: 14237
Ion Ratio Lower Upper
85 100
87 30.8 15.8 47.4



Abundance Scan 33 (1.288 min): VX031645.D\data.ms (-29)



#3

Chloromethane

Concen: 23.408 ug/l

RT: 1.288 min Scan# 31

Delta R.T. 0.000 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17

Instrument :

MSVOA_X

ClientSampleId :


VX0926WBS01

Tgt Ion: 50 Resp: 14396




Ion Ratio Lower Upper

50 100

52 36.7 26.9 40.3



Abundance Scan 47 (1.374 min): VX031645.D\data.ms (-42)



#4

Vinyl Chloride

Concen: 21.743 ug/l

RT: 1.368 min Scan# 46

Delta R.T. -0.006 min

Lab File: VX031651.D


Acq: 26 Sep 2022 16:17


Tgt Ion: 62 Resp: 19457

Ion Ratio Lower Upper

62 100

64 31.7 25.3 37.9









#5
Bromomethane
Concen: 27.483 ug/l
RT: 1.612 min Scan# 81
Delta R.T. 0.001 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion: 94 Resp: 24164
Ion Ratio Lower Upper
94 100
96 91.4 74.4 111.6



#6
Chloroethane
Concen: 21.771 ug/l
RT: 1.685 min Scan# 98
Delta R.T. -0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 64 Resp: 23881
Ion Ratio Lower Upper
64 100
66 31.4 25.0 37.4









#7
Trichlorofluoromethane
Concen: 20.341 ug/l
RT: 1.886 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion:101 Resp: 49309
Ion Ratio Lower Upper
101 100
103 62.1 51.8 77.6



#8
Diethyl Ether
Concen: 23.624 ug/l
RT: 2.130 min Scan# 171
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 74 Resp: 16078
Ion Ratio Lower Upper
74 100
45 77.3 48.5 145.6









#9
1,1,2-Trichlorotrifluoroethane
Concen: 20.748 ug/l
RT: 2.331 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion:101 Resp: 14735
Ion Ratio Lower Upper
101 100
85 47.8 38.6 57.8
151 77.1 54.0 81.0



#10
Methyl Iodide
Concen: 18.032 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion:142 Resp: 12553
Ion Ratio Lower Upper
142 100
127 47.2 37.4 56.0
141 14.6 11.6 17.4





#11

Tert butyl alcohol

Concen: 89.335 ug/l

RT: 2.995 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17

Instrument :

MSVOA_X

ClientSampleId :





VX0926WBS01

Tgt Ion: 59 Resp: 15726

Ion Ratio Lower Upper

59 100

57 9.9 8.2 12.2



#12

1,1-Dichloroethene

Concen: 22.428 ug/l

RT: 2.313 min Scan# 201

Delta R.T. -0.006 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17




Tgt Ion: 96 Resp: 13071

Ion Ratio Lower Upper


96 100

61 149.7 135.3 202.9


98 62.0 49.4 74.2




Abundance Scan 189 (2.239 min): VX031645.D\data.ms (-18



m/z--> Scan 189 (2.239 min): VX031651.D\data.ms



m/z--> Abundance Scan 189 (2.239 min): VX031651.D\data.ms (-14



m/z-->

#13

Acrolein

Concen: 104.842 ug/l

RT: 2.239 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17

Instrument :

MSVOA_X

ClientSampleId :

VX0926WBS01


Tgt Ion: 56 Resp: 8224

Ion Ratio Lower Upper


56 100

55 72.2 56.0 84.0


Abundance




Abundance Scan 259 (2.666 min): VX031645.D\data.ms (-25



m/z--> Scan 258 (2.660 min): VX031651.D\data.ms



m/z--> Abundance Scan 258 (2.660 min): VX031651.D\data.ms (-21



m/z-->

#14

Allyl chloride

Concen: 19.665 ug/l

RT: 2.660 min Scan# 258

Delta R.T. -0.006 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17

Tgt Ion: 41 Resp: 18375


Ion Ratio Lower Upper


41 100

39 79.8 58.2 87.4

76 37.9 26.9 40.3

Abundance









#15
Acrylonitrile
Concen: 99.554 ug/l
RT: 3.069 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion: 53 Resp: 39179
Ion Ratio Lower Upper
53 100
52 83.9 66.9 100.3
51 38.2 30.0 45.0



#16
Acetone
Concen: 112.985 ug/l
RT: 2.392 min Scan# 214
Delta R.T. 0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 43 Resp: 32401
Ion Ratio Lower Upper
43 100
58 34.4 23.8 35.8









#17
Carbon Disulfide
Concen: 20.938 ug/l
RT: 2.508 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion: 76 Resp: 25982
Ion Ratio Lower Upper
76 100
78 8.4 7.2 10.8



#18
Methyl Acetate
Concen: 20.472 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 43 Resp: 20861
Ion Ratio Lower Upper
43 100
74 26.2 18.6 27.8









#19
Methyl tert-butyl Ether
Concen: 19.514 ug/l
RT: 3.111 min Scan# 311
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion: 73 Resp: 49216
Ion Ratio Lower Upper
73 100
57 19.6 17.2 25.8



#20
Methylene Chloride
Concen: 21.643 ug/l
RT: 2.788 min Scan# 279
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 84 Resp: 16208
Ion Ratio Lower Upper
84 100
49 108.2 105.9 158.9
51 32.3 32.2 48.2
86 58.4 50.8 76.2









#21
trans-1,2-Dichloroethene
Concen: 19.306 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion: 96 Resp: 15158
Ion Ratio Lower Upper
96 100
61 110.0 106.7 160.1
98 65.3 48.7 73.1



#22
Diisopropyl ether
Concen: 19.711 ug/l
RT: 3.758 min Scan# 438
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 45 Resp: 47403
Ion Ratio Lower Upper
45 100
43 87.8 59.0 88.6
87 31.5 22.6 33.8
59 12.8 8.8 13.2









#23
 Vinyl Acetate
 Concen: 102.230 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. -0.006 min
 Lab File: VX031651.D
 Acq: 26 Sep 2022 16:17




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0926WBS01


Tgt Ion: 43 Resp: 194432
 Ion Ratio Lower Upper
 43 100
 86 12.6 8.7 13.1



#24
 1,1-Dichloroethane
 Concen: 19.468 ug/l
 RT: 3.605 min Scan# 413
 Delta R.T. -0.006 min
 Lab File: VX031651.D
 Acq: 26 Sep 2022 16:17

Tgt Ion: 63 Resp: 26991
 Ion Ratio Lower Upper
 63 100
 98 9.2 3.5 10.4
 100 5.7 2.4 7.2





#25

2-Butanone

Concen: 100.756 ug/l

RT: 4.562 min Scan# 5

Delta R.T. 0.000 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17

Instrument :

MSVOA_X

ClientSampleId :





VX0926WBS01

Tgt Ion: 43 Resp: 54536

Ion Ratio Lower Upper

43 100

72 29.7 21.1 31.7



#26

2,2-Dichloropropane

Concen: 18.272 ug/l

RT: 4.477 min Scan# 556

Delta R.T. -0.000 min

Lab File: VX031651.D




Acq: 26 Sep 2022 16:17


Tgt Ion: 77 Resp: 20561

Ion Ratio Lower Upper

77 100

97 24.2 10.9 32.6









#27
 cis-1,2-Dichloroethene
 Concen: 19.711 ug/l
 RT: 4.489 min Scan# 511
 Delta R.T. 0.000 min
 Lab File: VX031651.D
 Acq: 26 Sep 2022 16:17




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0926WBS01


Tgt Ion:	96	Resp:	18953
Ion	Ratio	Lower	Upper
96	100		
61	131.3	0.0	305.2
98	64.9	0.0	128.4



#28
 Bromochloromethane
 Concen: 20.611 ug/l
 RT: 4.898 min Scan# 625
 Delta R.T. -0.006 min
 Lab File: VX031651.D
 Acq: 26 Sep 2022 16:17

Tgt Ion:	49	Resp:	13368
Ion	Ratio	Lower	Upper
49	100		
129	1.5	0.0	3.6
130	85.5	59.0	88.4









#29
Tetrahydrofuran
Concen: 98.778 ug/l
RT: 5.007 min Scan# 644
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion:	42	Resp:	33587
Ion	Ratio	Lower	Upper
42	100		
72	52.7	36.4	54.6
71	50.2	33.5	50.3



#30
Chloroform
Concen: 19.272 ug/l
RT: 5.093 min Scan# 657
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion:	83	Resp:	32871
Ion	Ratio	Lower	Upper
83	100		
85	62.2	52.8	79.2









#31
Cyclohexane
Concen: 20.272 ug/l
RT: 5.477 min Scan# 719
Delta R.T. 0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion:	56	Resp:	21857
Ion	Ratio	Lower	Upper
56	100		
69	31.3	25.1	37.7
84	101.6	70.2	105.4



#32
1,1,1-Trichloroethane
Concen: 19.372 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion:	97	Resp:	27365
Ion	Ratio	Lower	Upper
97	100		
99	64.5	51.3	76.9
61	46.2	35.9	53.9





#33

1,2-Dichloroethane-d4

Concen: 48.231 ug/l

RT: 5.952 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17

Instrument :

MSVOA_X

ClientSampleId :





VX0926WBS01

Tgt Ion: 65 Resp: 54692

Ion Ratio Lower Upper

65 100

67 51.0 0.0 102.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.000 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17




Tgt Ion: 114 Resp: 79195


Ion Ratio Lower Upper

114 100

63 18.3 0.0 42.6


88 13.4 0.0 33.0








#35
Dibromofluoromethane
Concen: 49.975 ug/l
RT: 5.379 min Scan# 705
Delta R.T. -0.012 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01







Tgt Ion:113 Resp: 47158
Ion Ratio Lower Upper
113 100
111 99.9 82.4 123.6
192 19.1 14.1 21.1



#36
1,1-Dichloropropene
Concen: 20.083 ug/l
RT: 5.690 min Scan# 755
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 75 Resp: 21764
Ion Ratio Lower Upper
75 100
110 35.7 17.5 52.5
77 30.4 24.9 37.3





#37

Ethyl Acetate

Concen: 20.538 ug/l

RT: 4.709 min Scan# 596

Delta R.T. -0.012 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17

Instrument :

MSVOA_X

ClientSampleId :

VX0926WBS01





Tgt Ion: 43 Resp: 22450

Ion Ratio Lower Upper

43 100

61 14.2 11.1 16.7

70 10.0 8.6 13.0



#38

Carbon Tetrachloride

Concen: 19.874 ug/l

RT: 5.672 min Scan# 752

Delta R.T. -0.006 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17




Tgt Ion: 117 Resp: 24272


Ion Ratio Lower Upper

117 100

119 100.9 74.6 112.0

121 31.4 25.3 37.9









#39
Methylcyclohexane
Concen: 19.375 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion:	83	Resp:	23923
Ion	Ratio	Lower	Upper
83	100		
55	67.7	61.0	91.4
98	53.9	37.7	56.5



#40
Benzene
Concen: 19.744 ug/l
RT: 6.038 min Scan# 812
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion:	78	Resp:	65714
Ion	Ratio	Lower	Upper
78	100		
77	23.7	18.7	28.1









#41
Methacrylonitrile
Concen: 20.171 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01

Tgt Ion: 41 Resp: 11219




Ion	Ratio	Lower	Upper
41	100		
39	61.0	45.0	67.6
67	82.8	58.2	87.4
52	36.5	25.0	37.6




#42
1,2-Dichloroethane
Concen: 20.208 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 62 Resp: 26012

Ion	Ratio	Lower	Upper
62	100		
98	9.6	0.0	17.8









#43
Isopropyl Acetate
Concen: 20.468 ug/l
RT: 6.336 min Scan# 80
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion: 43 Resp: 35212
Ion Ratio Lower Upper
43 100
61 18.8 15.3 22.9
87 14.2 10.6 15.8



#44
Trichloroethene
Concen: 20.159 ug/l
RT: 7.123 min Scan# 990
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 130 Resp: 18951
Ion Ratio Lower Upper
130 100
95 91.1 0.0 206.8





#45

1,2-Dichloropropane

Concen: 20.661 ug/l

RT: 7.428 min Scan# 1040

Delta R.T. -0.006 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17

Instrument :

MSVOA_X

ClientSampleId :





VX0926WBS01

Tgt Ion: 63 Resp: 16689

Ion Ratio Lower Upper

63 100

65 31.8 26.6 39.8



#46

Dibromomethane

Concen: 21.230 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17




Tgt Ion: 93 Resp: 13772


Ion Ratio Lower Upper

93 100

95 81.2 65.5 98.3

174 93.7 67.6 101.4





#47

Bromodichloromethane

Concen: 20.952 ug/l

RT: 7.818 min Scan# 1104

Delta R.T. -0.006 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17

Instrument :

MSVOA_X

ClientSampleId :

VX0926WBS01





Tgt Ion: 83 Resp: 24527

Ion Ratio Lower Upper

83 100

85 60.0 51.0 76.6

127 8.6 7.0 10.6



#48

Methyl methacrylate

Concen: 21.742 ug/l

RT: 7.696 min Scan# 1084

Delta R.T. -0.000 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17




Tgt Ion: 41 Resp: 16450


Ion Ratio Lower Upper

41 100

69 90.8 64.6 96.8

39 77.5 51.4 77.0#





#49

1,4-Dioxane

Concen: 448.324 ug/l

RT: 7.690 min Scan# 1086

Delta R.T. 0.006 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17

Instrument :

MSVOA_X

ClientSampleId :

VX0926WBS01





Tgt Ion: 88 Resp: 8673

Ion Ratio Lower Upper

88 100

43 31.3 27.1 40.7

58 59.0 53.9 80.9



#50

Toluene-d8

Concen: 52.423 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.006 min

Lab File: VX031651.D




Acq: 26 Sep 2022 16:17


Tgt Ion: 98 Resp: 179342

Ion Ratio Lower Upper

98 100

100 66.4 52.5 78.7









#51
4-Methyl-2-Pentanone
Concen: 109.756 ug/l
RT: 8.574 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion: 43 Resp: 119518
Ion Ratio Lower Upper
43 100
58 40.2 30.6 46.0



#52
Toluene
Concen: 21.351 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 92 Resp: 47370
Ion Ratio Lower Upper
92 100
91 174.8 137.1 205.7





#53

t-1,3-Dichloropropene

Concen: 18.884 ug/l

RT: 8.976 min Scan# 1194

Delta R.T. -0.006 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17

Instrument :

MSVOA_X

ClientSampleId :





VX0926WBS01

Tgt Ion: 75 Resp: 23884

Ion Ratio Lower Upper

75 100

77 34.3 24.8 37.2



#54

cis-1,3-Dichloropropene

Concen: 20.314 ug/l

RT: 8.366 min Scan# 1194

Delta R.T. 0.000 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17




Tgt Ion: 75 Resp: 24713


Ion Ratio Lower Upper

75 100

77 30.5 24.4 36.6

39 54.1 43.7 65.5









#55
1,1,2-Trichloroethane
Concen: 20.308 ug/l
RT: 9.153 min Scan# 1323
Delta R.T. -0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01

Tgt Ion: 97 Resp: 19398




Ion	Ratio	Lower	Upper
97	100		
83	81.1	69.0	103.6
85	52.1	44.1	66.1
99	64.9	51.3	76.9




#56
Ethyl methacrylate
Concen: 22.166 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 69 Resp: 27090

Ion	Ratio	Lower	Upper
69	100		
41	61.9	54.2	81.4
39	44.7	34.2	51.2





#57

1,3-Dichloropropane

Concen: 20.421 ug/l

RT: 9.311 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17

Instrument :

MSVOA_X

ClientSampleId :





VX0926WBS01

Tgt Ion: 76 Resp: 30892

Ion Ratio Lower Upper

76 100

78 33.1 25.5 38.3



#58

2-Chloroethyl Vinyl ether

Concen: 106.075 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.006 min

Lab File: VX031651.D




Acq: 26 Sep 2022 16:17


Tgt Ion: 63 Resp: 63835

Ion Ratio Lower Upper

63 100

106 33.1 22.6 34.0





#59

2-Hexanone

Concen: 116.808 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. 0.000 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17

Instrument :

MSVOA_X

ClientSampleId :





VX0926WBS01

Tgt Ion: 43 Resp: 94903

Ion Ratio Lower Upper

43 100

58 53.3 26.7 80.0



#60

Dibromochloromethane

Concen: 18.401 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.006 min

Lab File: VX031651.D




Acq: 26 Sep 2022 16:17


Tgt Ion: 129 Resp: 19403

Ion Ratio Lower Upper

129 100

127 81.2 38.7 116.1









#61
1,2-Dibromoethane
Concen: 21.801 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion:107 Resp: 21139
Ion Ratio Lower Upper
107 100
109 92.7 74.2 111.4



#62
4-Bromofluorobenzene
Concen: 54.810 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 95 Resp: 68724
Ion Ratio Lower Upper
95 100
174 86.1 0.0 153.2
176 81.3 0.0 144.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion:117 Resp: 95021
Ion Ratio Lower Upper
117 100
82 50.7 46.8 70.2
119 30.8 26.4 39.6



#64
Tetrachloroethene
Concen: 19.549 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion:164 Resp: 19855
Ion Ratio Lower Upper
164 100
166 127.5 99.0 148.4
129 87.8 78.0 117.0
131 87.8 74.2 111.2









#65
Chlorobenzene
Concen: 20.533 ug/l
RT: 10.080 min Scan# 1475
Delta R.T. 0.001 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion:112 Resp: 54356
Ion Ratio Lower Upper
112 100
114 29.0 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 20.462 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion:131 Resp: 19490
Ion Ratio Lower Upper
131 100
133 95.5 46.6 139.7
119 62.4 31.9 95.9









#67
Ethyl Benzene
Concen: 20.438 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion: 91 Resp: 92036
Ion Ratio Lower Upper
91 100
106 32.3 25.2 37.8



#68
m/p-Xylenes
Concen: 42.534 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 106 Resp: 76960
Ion Ratio Lower Upper
106 100
91 200.4 161.0 241.4









#69
o-Xylene
Concen: 21.000 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01

Tgt Ion:106 Resp: 37140
Ion Ratio Lower Upper
106 100
91 207.2 107.2 321.6



#70
Styrene
Concen: 21.829 ug/l
RT: 10.659 min Scan# 1570
Delta R.T. -0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion:104 Resp: 62841
Ion Ratio Lower Upper
104 100
78 48.8 41.4 62.2
103 56.1 43.8 65.8









#71
Bromoform
Concen: 18.632 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion:173 Resp: 14058
Ion Ratio Lower Upper
173 100
175 49.5 24.0 72.0
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion:152 Resp: 78995
Ion Ratio Lower Upper
152 100
115 68.1 43.4 130.1
150 163.4 0.0 346.8









#73
Isopropylbenzene
Concen: 19.078 ug/l
RT: 10.964 min Scan# 1
Delta R.T. 0.001 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion:105 Resp: 97464
Ion Ratio Lower Upper
105 100
120 27.5 13.7 41.0



#74
N-ethyl acetate
Concen: 20.228 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 43 Resp: 33521
Ion Ratio Lower Upper
43 100
70 45.5 33.4 50.2
55 27.2 18.2 27.2
61 24.7 19.3 28.9









#75
1,1,2,2-Tetrachloroethane
Concen: 19.698 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion: 83 Resp: 31974
Ion Ratio Lower Upper
83 100
131 11.4 5.1 15.4
85 64.8 32.6 97.8



#76
1,2,3-Trichloropropane
Concen: 22.561 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 75 Resp: 32309
Ion Ratio Lower Upper
75 100
77 32.9 20.3 60.9









#77
Bromobenzene
Concen: 18.923 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion:156 Resp: 24575
Ion Ratio Lower Upper
156 100
77 149.2 79.5 238.6
158 95.4 48.4 145.3



#78
n-propylbenzene
Concen: 20.463 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 91 Resp: 116326
Ion Ratio Lower Upper
91 100
120 23.7 11.8 35.4








#79
2-Chlorotoluene
Concen: 19.997 ug/l
RT: 11.366 min Scan# 1100
Delta R.T. -0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion: 91 Resp: 70198
Ion Ratio Lower Upper
91 100
126 35.0 17.4 52.3



#80
1,3,5-Trimethylbenzene
Concen: 20.434 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 105 Resp: 88489
Ion Ratio Lower Upper
105 100
120 48.7 24.5 73.5









#81
trans-1,4-Dichloro-2-butene
Concen: 18.423 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion: 75 Resp: 6891
Ion Ratio Lower Upper
75 100
53 125.6 77.6 116.4#
89 44.7 37.4 56.0



#82
4-Chlorotoluene
Concen: 20.893 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 91 Resp: 84732
Ion Ratio Lower Upper
91 100
126 30.6 15.1 45.3









#83
tert-Butylbenzene
Concen: 19.858 ug/l
RT: 11.713 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion:119 Resp: 82928
Ion Ratio Lower Upper
119 100
91 61.6 29.2 87.6
134 25.6 11.6 34.8



#84
1,2,4-Trimethylbenzene
Concen: 20.836 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion:105 Resp: 89314
Ion Ratio Lower Upper
105 100
120 45.5 22.6 67.8









#85
 sec-Butylbenzene
 Concen: 20.683 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX031651.D
 Acq: 26 Sep 2022 16:17




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0926WBS01


Tgt Ion:105 Resp: 109771
 Ion Ratio Lower Upper
 105 100
 134 20.8 10.4 31.1



#86
 p-Isopropyltoluene
 Concen: 20.611 ug/l
 RT: 12.012 min Scan# 1792
 Delta R.T. 0.000 min
 Lab File: VX031651.D
 Acq: 26 Sep 2022 16:17

Tgt Ion:119 Resp: 94332
 Ion Ratio Lower Upper
 119 100
 134 26.5 13.4 40.2
 91 25.2 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 20.032 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. 0.000 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17

Instrument :

MSVOA_X

ClientSampleId :

VX0926WBS01





Tgt Ion:146 Resp: 50433

Ion Ratio Lower Upper

146 100

111 38.4 19.8 59.4

148 62.5 31.5 94.5



#88

1,4-Dichlorobenzene

Concen: 20.155 ug/l

RT: 12.043 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX031651.D

Acq: 26 Sep 2022 16:17




Tgt Ion:146 Resp: 53115


Ion Ratio Lower Upper

146 100

111 39.1 20.5 61.5

148 62.4 32.0 96.2









#89
n-Butylbenzene
Concen: 21.015 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01

Tgt Ion: 91 Resp: 78569




Ion	Ratio	Lower	Upper
91	100		
92	53.4	27.0	81.0
134	26.7	13.3	39.9




#90
Hexachloroethane
Concen: 19.932 ug/l
RT: 12.543 min Scan# 1879
Delta R.T. 0.001 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion: 117 Resp: 14628

Ion	Ratio	Lower	Upper
117	100		
201	65.9	34.8	104.4









#91
1,2-Dichlorobenzene
Concen: 20.568 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion:	146	Resp:	51858
Ion	Ratio	Lower	Upper
146	100		
111	39.4	20.6	61.8
148	61.8	31.6	95.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 19.479 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion:	75	Resp:	6336
Ion	Ratio	Lower	Upper
75	100		
155	85.8	40.3	120.8
157	112.9	51.5	154.5









#93
1,2,4-Trichlorobenzene
Concen: 19.375 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17




Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01


Tgt Ion:180 Resp: 28567
Ion Ratio Lower Upper
180 100
182 94.7 46.9 140.6
145 31.5 16.7 50.1



#94
Hexachlorobutadiene
Concen: 19.305 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion:225 Resp: 11099
Ion Ratio Lower Upper
225 100
223 61.2 30.8 92.4
227 63.8 32.3 96.9









#95
Naphthalene
Concen: 20.006 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Instrument :
MSVOA_X
ClientSampleId :
VX0926WBS01

Tgt Ion:	128	Resp:	98530
Ion Ratio	Lower	Upper	
128	100		
127	13.0	10.4	15.6
129	10.9	8.8	13.2



#96
1,2,3-Trichlorobenzene
Concen: 19.346 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. 0.000 min
Lab File: VX031651.D
Acq: 26 Sep 2022 16:17

Tgt Ion:	180	Resp:	29275
Ion Ratio	Lower	Upper	
180	100		
182	91.8	47.9	143.8
145	31.9	17.6	52.8

