

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031669.D
 Acq On : 27 Sep 2022 00:02
 Operator : JC/MD
 Sample : N4791-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-EB-20220922

Quant Time: Sep 27 01:57:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.556	168	47783	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	81161	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	89447	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	53761	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54668	53.905	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 107.820%			
35) Dibromofluoromethane	5.385	113	43145	44.615	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 89.240%			
50) Toluene-d8	8.647	98	171090	48.800	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 97.600%			
62) 4-Bromofluorobenzene	11.079	95	50989	39.680	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 79.360%#			
Target Compounds						Qvalue
3) Chloromethane	1.288	50	46	Below Cal	#	41
5) Bromomethane	1.618	94	458	Below Cal		83
6) Chloroethane	1.673	64	249	Below Cal	#	43
7) Trichlorofluoromethane	1.892	101	61	Below Cal	#	18
10) Methyl Iodide	2.459	142	228	0.366 ug/l	#	81
11) Tert butyl alcohol	2.947	59	142	Below Cal	#	73
16) Acetone	2.392	43	248	Below Cal	#	45
17) Carbon Disulfide	2.508	76	222	Below Cal	#	75
20) Methylene Chloride	2.794	84	185	Below Cal		89
28) Bromochloromethane	5.105	49	152	0.262 ug/l	#	15
30) Chloroform	5.099	83	2091	1.371 ug/l	#	77
31) Cyclohexane	5.544	56	1085	1.125 ug/l	#	29
36) 1,1-Dichloropropene	5.550	75	3638	3.276 ug/l	#	51
38) Carbon Tetrachloride	5.562	117	5031	4.020 ug/l	#	19
51) 4-Methyl-2-Pentanone	8.647	43	742	0.665 ug/l	#	1
76) 1,2,3-Trichloropropane	11.079	75	29233	29.994 ug/l	#	36
81) trans-1,4-Dichloro-2-b...	11.079	75	29233	98.864 ug/l	#	11


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
Data File : VX031669.D
Acq On : 27 Sep 2022 00:02
Operator : JC/MD
Sample : N4791-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-EB-20220922

Quant Time: Sep 27 01:57:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX031669.D
Acq: 27 Sep 2022 00:02




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-EB-20220922

Tgt Ion:168 Resp: 47783
Ion Ratio Lower Upper
168 100
99 50.9 48.8 73.2




#3
Chloromethane
Concen: Below Cal
RT: 1.288 min Scan# 33
Delta R.T. 0.000 min
Lab File: VX031669.D
Acq: 27 Sep 2022 00:02


Tgt Ion: 50 Resp: 46
Ion Ratio Lower Upper
50 100
52 0.0 26.9 40.3#




Abundance Scan 86 (1.611 min): VX031645.D\data.ms (-81)




Abundance Scan 87 (1.618 min): VX031669.D\data.ms




Abundance Scan 87 (1.618 min): VX031669.D\data.ms (-37)




Abundance Scan 98 (1.685 min): VX031645.D\data.ms (-93)




Abundance Scan 96 (1.673 min): VX031669.D\data.ms



Abundance Scan 96 (1.673 min): VX031669.D\data.ms (-49)



Abundance Scan 96 (1.673 min): VX031669.D\data.ms (-49)



#5

Bromomethane

Concen: Below Cal

RT: 1.618 min Scan# 81

Delta R.T. 0.007 min

Lab File: VX031669.D

Acq: 27 Sep 2022 00:02

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-RW5-EB-20220922


Tgt Ion: 94 Resp: 458

Ion Ratio Lower Upper

94 100

96 109.4 74.4 111.6

Abundance



#6

Chloroethane

Concen: Below Cal

RT: 1.673 min Scan# 96

Delta R.T. -0.012 min

Lab File: VX031669.D

Acq: 27 Sep 2022 00:02


Tgt Ion: 64 Resp: 249


Ion Ratio Lower Upper

64 100

66 0.0 25.0 37.4#

Abundance









#7
Trichlorofluoromethane
Concen: Below Cal
RT: 1.892 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX031669.D
Acq: 27 Sep 2022 00:02




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-EB-20220922


Tgt Ion:101 Resp: 61
Ion Ratio Lower Upper
101 100
103 0.0 51.8 77.6#



#10
Methyl Iodide
Concen: 0.366 ug/l
RT: 2.459 min Scan# 225
Delta R.T. 0.006 min
Lab File: VX031669.D
Acq: 27 Sep 2022 00:02

Tgt Ion:142 Resp: 228
Ion Ratio Lower Upper
142 100
127 37.3 37.4 56.0#
141 0.0 11.6 17.4#





#11

Tert butyl alcohol

Concen: Below Cal

RT: 2.947 min Scan# 30

Delta R.T. -0.042 min

Lab File: VX031669.D

Acq: 27 Sep 2022 00:02

Instrument :

MSVOA_X

ClientSampleId :





BP-VPB-RW5-EB-20220922

Tgt Ion: 59 Resp: 142

Ion Ratio Lower Upper

59 100

57 0.0 8.2 12.2#



#16

Acetone

Concen: Below Cal

RT: 2.392 min Scan# 214

Delta R.T. 0.006 min

Lab File: VX031669.D




Acq: 27 Sep 2022 00:02


Tgt Ion: 43 Resp: 248

Ion Ratio Lower Upper

43 100

58 0.0 23.8 35.8#









#17
Carbon Disulfide
Concen: Below Cal
RT: 2.508 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX031669.D
Acq: 27 Sep 2022 00:02




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-EB-20220922


Tgt Ion: 76 Resp: 222
Ion Ratio Lower Upper
76 100
78 0.0 7.2 10.8#



#20
Methylene Chloride
Concen: Below Cal
RT: 2.794 min Scan# 280
Delta R.T. 0.006 min
Lab File: VX031669.D
Acq: 27 Sep 2022 00:02

Tgt Ion: 84 Resp: 185
Ion Ratio Lower Upper
84 100
49 152.8 105.9 158.9
51 39.6 32.2 48.2
86 59.0 50.8 76.2









#28
Bromochloromethane
Concen: 0.262 ug/l
RT: 5.105 min Scan# 61
Delta R.T. 0.201 min
Lab File: VX031669.D
Acq: 27 Sep 2022 00:02




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-EB-20220922


Tgt Ion: 49 Resp: 152
Ion Ratio Lower Upper
49 100
129 0.0 0.0 3.6
130 0.0 59.0 88.4#



#30
Chloroform
Concen: 1.371 ug/l
RT: 5.099 min Scan# 658
Delta R.T. -0.000 min
Lab File: VX031669.D
Acq: 27 Sep 2022 00:02


Tgt Ion: 83 Resp: 2091
Ion Ratio Lower Upper
83 100
85 47.9 52.8 79.2#








#31
Cyclohexane
Concen: 1.125 ug/l
RT: 5.544 min Scan# 719
Delta R.T. 0.073 min
Lab File: VX031669.D
Acq: 27 Sep 2022 00:02

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-EB-20220922







Tgt Ion: 56 Resp: 1085
Ion Ratio Lower Upper
56 100
69 125.8 25.1 37.7#
84 55.1 70.2 105.4#



#33
1,2-Dichloroethane-d4
Concen: 53.905 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX031669.D
Acq: 27 Sep 2022 00:02

Tgt Ion: 65 Resp: 54668
Ion Ratio Lower Upper
65 100
67 49.6 0.0 102.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 91

Delta R.T. 0.000 min

Lab File: VX031669.D

Acq: 27 Sep 2022 00:02

Instrument :





MSVOA_X

ClientSampleId :

BP-VPB-RW5-EB-20220922

Tgt Ion:114 Resp: 81161

Ion	Ratio	Lower	Upper
114	100		
63	17.8	0.0	42.6
88	14.2	0.0	33.0



#35

Dibromofluoromethane

Concen: 44.615 ug/l

RT: 5.385 min Scan# 705




Delta R.T. -0.006 min


Lab File: VX031669.D

Acq: 27 Sep 2022 00:02

Tgt Ion:113 Resp: 43145

Ion	Ratio	Lower	Upper
113	100		
111	103.7	82.4	123.6
192	20.1	14.1	21.1





#36

1,1-Dichloropropene

Concen: 3.276 ug/l

RT: 5.550 min Scan# 71

Delta R.T. -0.140 min

Lab File: VX031669.D


Acq: 27 Sep 2022 00:02

Instrument :

MSVOA_X

ClientSampleId :

BP-VPB-RW5-EB-20220922






Tgt Ion: 75 Resp: 3638

Ion Ratio Lower Upper

75 100

110 10.1 17.5 52.5#

77 0.0 24.9 37.3#



#38

Carbon Tetrachloride

Concen: 4.020 ug/l

RT: 5.562 min Scan# 734

Delta R.T. -0.116 min

Lab File: VX031669.D

Acq: 27 Sep 2022 00:02




Tgt Ion: 117 Resp: 5031


Ion Ratio Lower Upper

117 100

119 7.9 74.6 112.0#

121 0.0 25.3 37.9#





#50

Toluene-d8

Concen: 48.800 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.006 min

Lab File: VX031669.D

Acq: 27 Sep 2022 00:02

Instrument :

MSVOA_X

ClientSampleId :





BP-VPB-RW5-EB-20220922

Tgt Ion: 98 Resp: 171090

Ion Ratio Lower Upper

98 100

100 63.5 52.5 78.7



#51

4-Methyl-2-Pentanone

Concen: 0.665 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. 0.067 min

Lab File: VX031669.D




Acq: 27 Sep 2022 00:02


Tgt Ion: 43 Resp: 742

Ion Ratio Lower Upper

43 100

58 185.7 30.6 46.0#









#62
4-Bromofluorobenzene
Concen: 39.680 ug/l
RT: 11.079 min Scan# 1
Delta R.T. 0.000 min
Lab File: VX031669.D
Acq: 27 Sep 2022 00:02




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-EB-20220922


Tgt Ion: 95 Resp: 50989
Ion Ratio Lower Upper
95 100
174 89.7 0.0 153.2
176 85.4 0.0 144.0



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX031669.D
Acq: 27 Sep 2022 00:02

Tgt Ion: 117 Resp: 89447
Ion Ratio Lower Upper
117 100
82 49.0 46.8 70.2
119 32.6 26.4 39.6









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX031669.D
Acq: 27 Sep 2022 00:02




Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-EB-20220922


Tgt Ion:152 Resp: 53761
Ion Ratio Lower Upper
152 100
115 57.3 43.4 130.1
150 159.7 0.0 346.8



#76
1,2,3-Trichloropropane
Concen: 29.994 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.159 min
Lab File: VX031669.D
Acq: 27 Sep 2022 00:02


Tgt Ion: 75 Resp: 29233
Ion Ratio Lower Upper
75 100
77 1.0 20.3 60.9#





#81
 trans-1,4-Dichloro-2-butene
 Concen: 98.864 ug/l
 RT: 11.079 min Scan# 1
 Delta R.T. 0.061 min
 Lab File: VX031669.D
 Acq: 27 Sep 2022 00:02

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-EB-20220922



Tgt Ion:	75	Resp:	29233
Ion	Ratio	Lower	Upper
75	100		
53	0.0	77.6	116.4#
89	0.0	37.4	56.0#

