

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031647.D
 Acq On : 26 Sep 2022 13:09
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Sep 27 01:52:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	96850	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	151825	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	156249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86038	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	89173	43.382	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	86.760%
35) Dibromofluoromethane	5.385	113	72395	40.019	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	80.040%
50) Toluene-d8	8.647	98	298809	45.561	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.120%
62) 4-Bromofluorobenzene	11.079	95	94005	39.107	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	78.220%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
Data File : VX031647.D
Acq On : 26 Sep 2022 13:09
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Sep 27 01:52:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

