

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031656.D
 Acq On : 26 Sep 2022 18:13
 Operator : JC/MD
 Sample : N4614-14DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP091322DL

Quant Time: Sep 27 01:54:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	33771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	68182	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	60770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	34741	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68875	96.092	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	192.180%#
35) Dibromofluoromethane	5.385	113	49620	61.078	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	122.160%
50) Toluene-d8	8.647	98	113681	38.597	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	77.200%#
62) 4-Bromofluorobenzene	11.079	95	39412	36.509	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	73.020%#
Target Compounds						
						Qvalue
19) Methyl tert-butyl Ether	3.117	73	7014	4.400	ug/l	# 44
25) 2-Butanone	4.574	43	5255	15.360	ug/l	# 85
31) Cyclohexane	5.458	56	4588	6.732	ug/l	# 80
39) Methylcyclohexane	7.379	83	2529	2.379	ug/l	86
40) Benzene	6.037	78	465143	162.325	ug/l	99
52) Toluene	8.720	92	147258	77.095	ug/l	99
67) Ethyl Benzene	10.195	91	143300	49.757	ug/l	99
68) m/p-Xylenes	10.299	106	234948	203.035	ug/l	98
69) o-Xylene	10.640	106	92249	81.559	ug/l	98
73) Isopropylbenzene	10.963	105	6777	3.016	ug/l	100
78) n-propylbenzene	11.305	91	18597	7.438	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	41859	21.979	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	162011	85.941	ug/l	99
95) Naphthalene	13.774	128	106009	48.943	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
Data File : VX031656.D
Acq On : 26 Sep 2022 18:13
Operator : JC/MD
Sample : N4614-14DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP091322DL

Quant Time: Sep 27 01:54:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

