

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031661.D
 Acq On : 26 Sep 2022 20:09
 Operator : JC/MD
 Sample : N4731-02DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-2DL

Quant Time: Sep 27 01:55:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	37031	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	66885	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	67408	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41008	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54140	68.885	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	137.760%#
35) Dibromofluoromethane	5.385	113	40961	51.397	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.800%
50) Toluene-d8	8.647	98	126148	43.661	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	87.320%
62) 4-Bromofluorobenzene	11.079	95	40785	38.514	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	77.020%#

Target Compounds					Qvalue	
20) Methylene Chloride	2.788	84	2062	2.542	ug/l	# 82
40) Benzene	6.038	78	293811	104.522	ug/l	99
43) Isopropyl Acetate	6.355	43	3655	2.516	ug/l	95
52) Toluene	8.720	92	124874	66.644	ug/l	99
67) Ethyl Benzene	10.195	91	11329	3.546	ug/l	93
68) m/p-Xylenes	10.299	106	41136	32.048	ug/l	98
69) o-Xylene	10.640	106	34592	27.572	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	9345	4.157	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	25667	11.535	ug/l	96
95) Naphthalene	13.780	128	1279680	500.520	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
Data File : VX031661.D
Acq On : 26 Sep 2022 20:09
Operator : JC/MD
Sample : N4731-02DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-2DL

Quant Time: Sep 27 01:55:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

