

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031672.D
 Acq On : 27 Sep 2022 01:11
 Operator : JC/MD
 Sample : N4791-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-GW-568-570

Quant Time: Sep 27 04:40:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 09/27/2022
 Supervised By :Mahesh Dadoda 09/28/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	51556	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	80916	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	92481	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58025	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	54997	50.261	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.520%			
35) Dibromofluoromethane	5.385	113	44273	45.920	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 91.840%			
50) Toluene-d8	8.653	98	171912	49.183	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 98.360%			
62) 4-Bromofluorobenzene	11.079	95	56027	43.733	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 87.460%			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	918	1.463	ug/l	98
9) 1,1,2-Trichlorotrifluo...	2.331	101	18805	27.440	ug/l	91
12) 1,1-Dichloroethene	2.318	96	7375	11.888	ug/l	90
27) cis-1,2-Dichloroethene	4.489	96	1887	2.034	ug/l	74
30) Chloroform	5.086	83	2006m	1.219	ug/l	
32) 1,1,1-Trichloroethane	5.373	97	805	0.591	ug/l #	47
38) Carbon Tetrachloride	5.678	117	2300	1.843	ug/l #	83
44) Trichloroethene	7.129	130	391138	407.212	ug/l	92
55) 1,1,2-Trichloroethane	9.159	97	795	0.815	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
Data File : VX031672.D
Acq On : 27 Sep 2022 01:11
Operator : JC/MD
Sample : N4791-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-GW-568-570

Quant Time: Sep 27 04:40:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 09/27/2022
Supervised By :Mahesh Dadoda 09/28/2022

