

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031673.D
 Acq On : 27 Sep 2022 01:35
 Operator : JC/MD
 Sample : N4791-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-RW5-GW-578-580

Quant Time: Sep 27 04:40:51 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	50332	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	79879	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	90358	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55824	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	55441	51.899	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 103.800%		
35) Dibromofluoromethane	5.391	113	45062	47.345	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 94.700%		
50) Toluene-d8	8.653	98	177312	51.386	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 102.780%		
62) 4-Bromofluorobenzene	11.079	95	54666	43.225	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 86.440%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	989	1.615	ug/l	92
9) 1,1,2-Trichlorotrifluo...	2.331	101	34436	51.470	ug/l	92
12) 1,1-Dichloroethene	2.319	96	11546	20.846	ug/l	85
24) 1,1-Dichloroethane	3.617	63	2577	1.973	ug/l #	91
27) cis-1,2-Dichloroethene	4.489	96	2548	2.813	ug/l	77
30) Chloroform	5.099	83	2565	1.596	ug/l	84
32) 1,1,1-Trichloroethane	5.385	97	1399	1.051	ug/l	94
38) Carbon Tetrachloride	5.672	117	491	0.399	ug/l #	72
44) Trichloroethene	7.129	130	655813	691.628	ug/l	93
55) 1,1,2-Trichloroethane	9.153	97	1065	1.105	ug/l #	86
64) Tetrachloroethene	9.275	164	740	0.766	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
Data File : VX031673.D
Acq On : 27 Sep 2022 01:35
Operator : JC/MD
Sample : N4791-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-RW5-GW-578-580

Quant Time: Sep 27 04:40:51 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

