

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031674.D
 Acq On : 27 Sep 2022 01:58
 Operator : JC/MD
 Sample : N4791-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Sep 27 04:41:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	47249	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	76244	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	83544	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	50836	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	54314	54.161	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.320%	
35) Dibromofluoromethane	5.385	113	44561	49.051	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.100%	
50) Toluene-d8	8.653	98	161003	48.884	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 97.760%	
62) 4-Bromofluorobenzene	11.079	95	50267	41.641	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 83.280%	
Target Compounds						
2) Dichlorodifluoromethane	1.166	85	385	0.670	ug/l	Qvalue 94
44) Trichloroethene	7.129	130	1747	1.930	ug/l	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
Data File : VX031674.D
Acq On : 27 Sep 2022 01:58
Operator : JC/MD
Sample : N4791-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Sep 27 04:41:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

