

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031680.D
 Acq On : 27 Sep 2022 04:17
 Operator : JC/MD
 Sample : N4764-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 355

Quant Time: Sep 27 05:20:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	50475	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	84620	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	106340	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	60925	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60937	56.882	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.760%
35) Dibromofluoromethane	5.385	113	48455	48.058	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.120%
50) Toluene-d8	8.647	98	183150	50.104	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.200%
62) 4-Bromofluorobenzene	11.079	95	63800	47.620	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.240%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	6491	7.747	ug/l	93
20) Methylene Chloride	2.788	84	5467	6.601	ug/l	94
37) Ethyl Acetate	4.721	43	9846	8.430	ug/l	97
43) Isopropyl Acetate	6.342	43	46974	25.554	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
Data File : VX031680.D
Acq On : 27 Sep 2022 04:17
Operator : JC/MD
Sample : N4764-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
355

Quant Time: Sep 27 05:20:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

