

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX092622\  
 Data File : VX031682.D  
 Acq On : 27 Sep 2022 05:03  
 Operator : JC/MD  
 Sample : N4793-01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 31 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Sep 27 08:15:30 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X092622W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Sep 27 01:44:18 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units      | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|------------|----------|
| -----                       |        |       |          |          |            |          |
| Internal Standards          |        |       |          |          |            |          |
| 1) Pentafluorobenzene       | 5.556  | 168   | 56838    | 50.000   | ug/l       | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763  | 114   | 89582    | 50.000   | ug/l       | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 97548    | 50.000   | ug/l       | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 60768    | 50.000   | ug/l       | 0.00     |
| System Monitoring Compounds |        |       |          |          |            |          |
| 33) 1,2-Dichloroethane-d4   | 5.958  | 65    | 61650    | 51.105   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | = 102.220% |          |
| 35) Dibromofluoromethane    | 5.391  | 113   | 49549    | 46.421   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | = 92.840%  |          |
| 50) Toluene-d8              | 8.647  | 98    | 187962   | 48.572   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | = 97.140%  |          |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 58905    | 41.531   | ug/l       | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | = 83.060%  |          |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX092622\  
Data File : VX031682.D  
Acq On : 27 Sep 2022 05:03  
Operator : JC/MD  
Sample : N4793-01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 31 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
STORAGE-BLANK-SOIL-REF-6

Quant Time: Sep 27 08:15:30 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X092622W.M  
Quant Title : SW846 8260  
QLast Update : Tue Sep 27 01:44:18 2022  
Response via : Initial Calibration

