

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031685.D
 Acq On : 27 Sep 2022 06:13
 Operator : JC/MD
 Sample : N4793-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Sep 27 08:16:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	51694	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	86203	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	94450	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	58480	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	58377	53.207	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.420%
35) Dibromofluoromethane	5.385	113	45896	44.684	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.360%
50) Toluene-d8	8.647	98	182713	49.067	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.140%
62) 4-Bromofluorobenzene	11.079	95	57217	41.923	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
Data File : VX031685.D
Acq On : 27 Sep 2022 06:13
Operator : JC/MD
Sample : N4793-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Sep 27 08:16:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

