

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX092622\
 Data File : VX031662.D
 Acq On : 26 Sep 2022 20:33
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 27 01:56:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	102	0.00
2 T	Dichlorodifluoromethane	50.000	45.511	9.0	100	0.00
3 P	Chloromethane	50.000	54.100	-8.2	105	0.00
4 C	Vinyl Chloride	50.000	46.171	7.7#	102	0.00
5 T	Bromomethane	50.000	67.992	-36.0#	125	0.00
6 T	Chloroethane	50.000	49.270	1.5	101	0.00
7 T	Trichlorofluoromethane	50.000	48.037	3.9	101	0.00
8 T	Diethyl Ether	50.000	57.094	-14.2	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	41.605	16.8	95	0.00
10 T	Methyl Iodide	50.000	49.459	1.1	95	0.00
11 T	Tert butyl alcohol	250.000	261.652	-4.7	100	0.00
12 CM	1,1-Dichloroethene	50.000	51.458	-2.9#	101	0.00
13 T	Acrolein	250.000	270.559	-8.2	102	0.00
14 T	Allyl chloride	50.000	44.484	11.0	97	0.00
15 T	Acrylonitrile	250.000	243.430	2.6	104	0.00
16 T	Acetone	250.000	286.186	-14.5	105	0.00
17 T	Carbon Disulfide	50.000	46.874	6.3	97	0.00
18 T	Methyl Acetate	50.000	45.775	8.5	102	0.00
19 T	Methyl tert-butyl Ether	50.000	51.451	-2.9	107	0.00
20 T	Methylene Chloride	50.000	49.371	1.3	102	0.00
21 T	trans-1,2-Dichloroethene	50.000	47.249	5.5	104	0.00
22 T	Diisopropyl ether	50.000	52.784	-5.6	102	0.00
23 T	Vinyl Acetate	250.000	271.317	-8.5	102	0.00
24 P	1,1-Dichloroethane	50.000	48.165	3.7	104	0.00
25 T	2-Butanone	250.000	262.031	-4.8	106	0.00
26 T	2,2-Dichloropropane	50.000	40.760	18.5	81	0.00
27 T	cis-1,2-Dichloroethene	50.000	49.047	1.9	103	0.00
28 T	Bromochloromethane	50.000	52.380	-4.8	102	0.00
29 T	Tetrahydrofuran	250.000	273.365	-9.3	105	0.00
30 C	Chloroform	50.000	51.796	-3.6#	103	0.00
31 T	Cyclohexane	50.000	50.262	-0.5	100	0.00
32 T	1,1,1-Trichloroethane	50.000	52.719	-5.4	101	0.00
33 S	1,2-Dichloroethane-d4	50.000	50.016	-0.0	103	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	105	0.00
35 S	Dibromofluoromethane	50.000	48.814	2.4	104	0.00
36 T	1,1-Dichloropropene	50.000	49.774	0.5	103	0.00
37 T	Ethyl Acetate	50.000	51.315	-2.6	102	0.00
38 T	Carbon Tetrachloride	50.000	50.405	-0.8	101	0.00
39 T	Methylcyclohexane	50.000	46.808	6.4	100	0.00
40 TM	Benzene	50.000	49.289	1.4	102	0.00
41 T	Methacrylonitrile	50.000	52.023	-4.0	104	0.00
42 TM	1,2-Dichloroethane	50.000	51.198	-2.4	101	0.00
43 T	Isopropyl Acetate	50.000	52.626	-5.3	101	0.00
44 TM	Trichloroethene	50.000	48.581	2.8	104	0.00
45 C	1,2-Dichloropropane	50.000	48.385	3.2#	101	0.00
46 T	Dibromomethane	50.000	50.183	-0.4	105	0.00
47 T	Bromodichloromethane	50.000	52.949	-5.9	103	0.00
48 T	Methyl methacrylate	50.000	55.479	-11.0	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031662.D
 Acq On : 26 Sep 2022 20:33
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 27 01:56:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	1142.468	-14.2	107	0.01
50 S	Toluene-d8	50.000	52.041	-4.1	103	0.00
51 T	4-Methyl-2-Pentanone	250.000	274.424	-9.8	103	0.00
52 CM	Toluene	50.000	54.350	-8.7#	104	0.00
53 T	t-1,3-Dichloropropene	50.000	48.448	3.1	98	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.773	-7.5	99	0.00
55 T	1,1,2-Trichloroethane	50.000	53.349	-6.7	103	0.00
56 T	Ethyl methacrylate	50.000	59.497	-19.0	105	0.00
57 T	1,3-Dichloropropane	50.000	53.083	-6.2	103	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	275.012	-10.0	105	0.00
59 T	2-Hexanone	250.000	278.830	-11.5	103	0.00
60 T	Dibromochloromethane	50.000	49.533	0.9	100	0.00
61 T	1,2-Dibromoethane	50.000	57.957	-15.9	103	0.00
62 S	4-Bromofluorobenzene	50.000	56.777	-13.6	102	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	100	0.00
64 T	Tetrachloroethene	50.000	47.048	5.9	100	0.00
65 PM	Chlorobenzene	50.000	51.214	-2.4	103	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	53.853	-7.7	102	0.00
67 C	Ethyl Benzene	50.000	51.640	-3.3#	102	0.00
68 T	m/p-Xylenes	100.000	105.229	-5.2	100	0.00
69 T	o-Xylene	50.000	53.888	-7.8	101	0.00
70 T	Styrene	50.000	55.848	-11.7	103	0.00
71 P	Bromoform	50.000	48.599	2.8	100	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	97	0.00
73 T	Isopropylbenzene	50.000	50.913	-1.8	101	0.00
74 T	N-ethyl acetate	50.000	53.188	-6.4	99	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.478	-1.0	100	0.00
76 T	1,2,3-Trichloropropane	50.000	50.823	-1.6	99	0.00
77 T	Bromobenzene	50.000	50.244	-0.5	104	0.00
78 T	n-propylbenzene	50.000	51.763	-3.5	98	0.00
79 T	2-Chlorotoluene	50.000	50.922	-1.8	101	0.00
80 T	1,3,5-Trimethylbenzene	50.000	53.131	-6.3	101	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.322	3.4	94	0.00
82 T	4-Chlorotoluene	50.000	52.084	-4.2	100	0.00
83 T	tert-Butylbenzene	50.000	54.662	-9.3	103	0.00
84 T	1,2,4-Trimethylbenzene	50.000	54.365	-8.7	102	0.00
85 T	sec-Butylbenzene	50.000	52.584	-5.2	100	0.00
86 T	p-Isopropyltoluene	50.000	52.550	-5.1	98	0.00
87 T	1,3-Dichlorobenzene	50.000	50.597	-1.2	99	0.00
88 T	1,4-Dichlorobenzene	50.000	50.010	-0.0	101	0.00
89 T	n-Butylbenzene	50.000	51.863	-3.7	97	0.00
90 T	Hexachloroethane	50.000	52.935	-5.9	97	0.00
91 T	1,2-Dichlorobenzene	50.000	51.784	-3.6	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	56.278	-12.6	98	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.744	-1.5	98	0.00
94 T	Hexachlorobutadiene	50.000	46.680	6.6	98	0.00
95 T	Naphthalene	50.000	56.989	-14.0	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
Data File : VX031662.D
Acq On : 26 Sep 2022 20:33
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 27 01:56:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.719	-1.4	99	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6