

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031686.D
 Acq On : 27 Sep 2022 06:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 27 08:16:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 26 12:56:44 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	104	0.00
2 T	Dichlorodifluoromethane	0.609	0.517	15.1	94	0.00
3 P	Chloromethane	0.667	0.569	14.7	104	0.00
4 C	Vinyl Chloride	0.837	0.744	11.1#	100	0.00
5 T	Bromomethane	0.931	0.857	7.9	110	0.00
6 T	Chloroethane	1.174	0.923	21.4	96	0.00
7 T	Trichlorofluoromethane	2.287	1.840	19.5	97	0.00
8 T	Diethyl Ether	0.723	0.605	16.3	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.665	0.570	14.3	99	0.00
10 T	Methyl Iodide	0.651	0.637	2.2	95	0.00
11 T	Tert butyl alcohol	0.167	0.135	19.2	96	0.00
12 CM	1,1-Dichloroethene	0.629	0.527	16.2#	103	0.00
13 T	Acrolein	0.080	0.066	17.5	99	0.00
14 T	Allyl chloride	0.874	0.781	10.6	99	0.00
15 T	Acrylonitrile	0.368	0.357	3.0	105	0.00
16 T	Acetone	0.309	0.268	13.3	102	0.00
17 T	Carbon Disulfide	1.303	1.047	19.6	98	0.00
18 T	Methyl Acetate	0.954	0.893	6.4	106	0.00
19 T	Methyl tert-butyl Ether	2.360	2.436	-3.2	109	0.00
20 T	Methylene Chloride	0.803	0.672	16.3	105	0.00
21 T	trans-1,2-Dichloroethene	0.735	0.700	4.8	106	0.00
22 T	Diisopropyl ether	2.251	2.232	0.8	97	0.00
23 T	Vinyl Acetate	1.780	1.803	-1.3	96	0.00
24 P	1,1-Dichloroethane	1.297	1.253	3.4	106	0.00
25 T	2-Butanone	0.507	0.493	2.8	100	0.00
26 T	2,2-Dichloropropane	1.053	0.702	33.3#	67	0.00
27 T	cis-1,2-Dichloroethene	0.900	0.884	1.8	105	0.00
28 T	Bromochloromethane	0.607	0.612	-0.8	100	0.00
29 T	Tetrahydrofuran	0.318	0.319	-0.3	97	0.00
30 C	Chloroform	1.596	1.614	-1.1#	102	0.00
31 T	Cyclohexane	1.009	1.020	-1.1	102	0.00
32 T	1,1,1-Trichloroethane	1.322	1.420	-7.4	104	0.00
33 S	1,2-Dichloroethane-d4	1.061	1.046	1.4	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	109	0.00
35 S	Dibromofluoromethane	0.596	0.559	6.2	104	0.00
36 T	1,1-Dichloropropene	0.684	0.651	4.8	102	0.00
37 T	Ethyl Acetate	0.690	0.636	7.8	96	0.00
38 T	Carbon Tetrachloride	0.771	0.780	-1.2	106	0.00
39 T	Methylcyclohexane	0.780	0.729	6.5	104	0.00
40 TM	Benzene	2.101	2.002	4.7	103	0.00
41 T	Methacrylonitrile	0.351	0.335	4.6	99	0.00
42 TM	1,2-Dichloroethane	0.813	0.790	2.8	99	0.00
43 T	Isopropyl Acetate	1.086	1.044	3.9	96	0.00
44 TM	Trichloroethene	0.594	0.576	3.0	108	0.00
45 C	1,2-Dichloropropane	0.510	0.488	4.3#	104	0.00
46 T	Dibromomethane	0.410	0.398	2.9	106	0.00
47 T	Bromodichloromethane	0.739	0.756	-2.3	104	0.00
48 T	Methyl methacrylate	0.478	0.487	-1.9	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031686.D
 Acq On : 27 Sep 2022 06:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 27 08:16:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Mon Sep 26 12:56:44 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.012	0.013	-8.3	108	0.01
50 S	Toluene-d8	2.160	2.121	1.8	101	0.00
51 T	4-Methyl-2-Pentanone	0.688	0.685	0.4	97	0.00
52 CM	Toluene	1.401	1.418	-1.2#	100	0.00
53 T	t-1,3-Dichloropropene	0.710	0.749	-5.5	92	0.00
54 T	cis-1,3-Dichloropropene	0.768	0.761	0.9	95	0.00
55 T	1,1,2-Trichloroethane	0.603	0.609	-1.0	101	0.00
56 T	Ethyl methacrylate	0.772	0.829	-7.4	99	0.00
57 T	1,3-Dichloropropane	0.955	0.948	0.7	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.380	0.391	-2.9	102	0.00
59 T	2-Hexanone	0.513	0.522	-1.8	98	0.00
60 T	Dibromochloromethane	0.606	0.659	-8.7	98	0.00
61 T	1,2-Dibromoethane	0.612	0.663	-8.3	100	0.00
62 S	4-Bromofluorobenzene	0.792	0.850	-7.3	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
64 T	Tetrachloroethene	0.534	0.521	2.4	107	0.00
65 PM	Chlorobenzene	1.393	1.356	2.7	100	0.00
66 T	1,1,1,2-Tetrachloroethane	0.501	0.516	-3.0	100	0.00
67 C	Ethyl Benzene	2.370	2.345	1.1#	100	0.00
68 T	m/p-Xylenes	0.952	0.968	-1.7	99	0.00
69 T	o-Xylene	0.931	0.954	-2.5	98	0.00
70 T	Styrene	1.515	1.609	-6.2	100	0.00
71 P	Bromoform	0.357	0.393	-10.1	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.234	3.166	2.1	100	0.00
74 T	N-amyl acetate	1.049	1.030	1.8	94	0.00
75 P	1,1,2,2-Tetrachloroethane	1.027	0.974	5.2	96	0.00
76 T	1,2,3-Trichloropropane	0.906	0.861	5.0	95	0.00
77 T	Bromobenzene	0.822	0.792	3.6	102	0.00
78 T	n-propylbenzene	3.598	3.589	0.3	97	0.00
79 T	2-Chlorotoluene	2.222	2.191	1.4	100	0.00
80 T	1,3,5-Trimethylbenzene	2.741	2.761	-0.7	98	0.00
81 T	trans-1,4-Dichloro-2-butene	0.227	0.211	7.0	79	0.00
82 T	4-Chlorotoluene	2.567	2.546	0.8	98	0.00
83 T	tert-Butylbenzene	2.643	2.721	-3.0	99	0.00
84 T	1,2,4-Trimethylbenzene	2.713	2.755	-1.5	98	0.00
85 T	sec-Butylbenzene	3.359	3.380	-0.6	98	0.00
86 T	p-Isopropyltoluene	2.897	2.931	-1.2	97	0.00
87 T	1,3-Dichlorobenzene	1.594	1.539	3.5	97	0.00
88 T	1,4-Dichlorobenzene	1.668	1.603	3.9	99	0.00
89 T	n-Butylbenzene	2.366	2.348	0.8	96	0.00
90 T	Hexachloroethane	0.465	0.482	-3.7	98	0.00
91 T	1,2-Dichlorobenzene	1.596	1.572	1.5	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.206	0.215	-4.4	94	0.00
93 T	1,2,4-Trichlorobenzene	0.933	0.904	3.1	96	0.00
94 T	Hexachlorobutadiene	0.364	0.339	6.9	101	0.00
95 T	Naphthalene	3.117	3.261	-4.6	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
Data File : VX031686.D
Acq On : 27 Sep 2022 06:36
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Sep 27 08:16:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Mon Sep 26 12:56:44 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.958	0.915	4.5	96	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6