

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
 Data File : VX031649.D
 Acq On : 26 Sep 2022 14:03
 Operator : JC/MD
 Sample : VX0926MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0926MBL01

Quant Time: Sep 27 01:53:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 27 01:44:18 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	55058	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	84549	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	89448	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59607	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	55984	47.909	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.820%
35) Dibromofluoromethane	5.385	113	46234	45.894	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.780%
50) Toluene-d8	8.647	98	167292	45.804	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	91.600%
62) 4-Bromofluorobenzene	11.079	95	55143	41.193	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	82.380%#


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX092622\
Data File : VX031649.D
Acq On : 26 Sep 2022 14:03
Operator : JC/MD
Sample : VX0926MBL01
Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0926MBL01

Quant Time: Sep 27 01:53:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X092622W.M
Quant Title : SW846 8260
QLast Update : Tue Sep 27 01:44:18 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 712
Delta R.T. -0.006 min
Lab File: VX031649.D
Acq: 26 Sep 2022 14:03




Instrument :
MSVOA_X
ClientSampleId :
VX0926MBL01


Tgt Ion:168 Resp: 55058
Ion Ratio Lower Upper
168 100
99 52.9 48.8 73.2



#33
1,2-Dichloroethane-d4
Concen: 47.909 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.006 min
Lab File: VX031649.D
Acq: 26 Sep 2022 14:03

Tgt Ion: 65 Resp: 55984
Ion Ratio Lower Upper
65 100
67 50.1 0.0 102.4








#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX031649.D
Acq: 26 Sep 2022 14:03




Instrument :
MSVOA_X
ClientSampleId :
VX0926MBL01


Tgt Ion:114 Resp: 84549
Ion Ratio Lower Upper
114 100
63 18.2 0.0 42.6
88 13.4 0.0 33.0



#35
Dibromofluoromethane
Concen: 45.894 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.006 min
Lab File: VX031649.D
Acq: 26 Sep 2022 14:03

Tgt Ion:113 Resp: 46234
Ion Ratio Lower Upper
113 100
111 101.7 82.4 123.6
192 19.8 14.1 21.1





#50

Toluene-d8

Concen: 45.804 ug/l

RT: 8.647 min Scan# 11

Delta R.T. -0.006 min

Lab File: VX031649.D

Acq: 26 Sep 2022 14:03

Instrument :

MSVOA_X

ClientSampleId :





VX0926MBL01

Tgt Ion: 98 Resp: 167292

Ion Ratio Lower Upper

98 100

100 65.2 52.5 78.7



#62

4-Bromofluorobenzene

Concen: 41.193 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.000 min

Lab File: VX031649.D

Acq: 26 Sep 2022 14:03




Tgt Ion: 95 Resp: 55143


Ion Ratio Lower Upper

95 100

174 86.3 0.0 153.2

176 83.1 0.0 144.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX031649.D
Acq: 26 Sep 2022 14:03

Instrument :
MSVOA_X
ClientSampleId :
VX0926MBL01

Tgt Ion:117 Resp: 89448
Ion Ratio Lower Upper
117 100
82 53.3 46.8 70.2
119 31.4 26.4 39.6



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX031649.D
Acq: 26 Sep 2022 14:03

Tgt Ion:152 Resp: 59607
Ion Ratio Lower Upper
152 100
115 56.6 43.4 130.1
150 158.6 0.0 346.8

