

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093025\
 Data File : VX047897.D
 Acq On : 30 Sep 2025 09:48
 Operator : JC/MD
 Sample : VX0930MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0930MBL01

Quant Time: Oct 01 01:04:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.532	168	107169	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.745	114	221661	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	224146	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	108964	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	95883	56.208	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.420%
35) Dibromofluoromethane	5.367	113	79611	53.553	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.100%
50) Toluene-d8	8.635	98	274975	53.457	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.920%
62) 4-Bromofluorobenzene	11.061	95	112606	57.682	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	115.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093025\
Data File : VX047897.D
Acq On : 30 Sep 2025 09:48
Operator : JC/MD
Sample : VX0930MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0930MBL01

Quant Time: Oct 01 01:04:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

