

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093025\
 Data File : VX047900.D
 Acq On : 30 Sep 2025 11:27
 Operator : JC/MD
 Sample : Q3213-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-184-GW-360-362

Quant Time: Oct 01 01:06:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	131080	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	265322	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	269226	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	129059	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	97917	46.930	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.860%
35) Dibromofluoromethane	5.373	113	84323	47.389	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.780%
50) Toluene-d8	8.635	98	291076	47.275	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.560%
62) 4-Bromofluorobenzene	11.067	95	120328	51.495	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.980%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	12510	13.790	ug/l	99
25) 2-Butanone	4.556	43	4247	3.753	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093025\
Data File : VX047900.D
Acq On : 30 Sep 2025 11:27
Operator : JC/MD
Sample : Q3213-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-184-GW-360-362

Quant Time: Oct 01 01:06:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

