

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093025\
 Data File : VX047903.D
 Acq On : 30 Sep 2025 12:30
 Operator : JC/MD
 Sample : Q3156-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB02-20250918

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/01/2025
 Supervised By : Mahesh Dadoda 10/03/2025

Quant Time: Oct 01 01:08:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	125275	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	259422	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	266390	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	130961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	102146	51.225	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.460%
35) Dibromofluoromethane	5.373	113	85252	49.000	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.000%
50) Toluene-d8	8.635	98	292362	48.564	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.120%
62) 4-Bromofluorobenzene	11.067	95	123949	54.250	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.500%
Target Compounds						Qvalue
16) Acetone	2.374	43	12160	14.025	ug/l	93
29) Tetrahydrofuran	5.001	42	3794m	5.831	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093025\
Data File : VX047903.D
Acq On : 30 Sep 2025 12:30
Operator : JC/MD
Sample : Q3156-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB02-20250918

Manual Integrations
APPROVED

Reviewed By : John Carlone 10/01/2025
Supervised By : Mahesh Dadoda 10/03/2025

Quant Time: Oct 01 01:08:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

