

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093025\
 Data File : VX047911.D
 Acq On : 30 Sep 2025 15:19
 Operator : JC/MD
 Sample : Q3193-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GM38-RW3-MW1-20250922

Quant Time: Oct 01 01:12:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	124889	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	257980	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	261260	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	132033	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	100743	50.678	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.360%
35) Dibromofluoromethane	5.379	113	84599	48.897	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.800%
50) Toluene-d8	8.635	98	291075	48.620	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.240%
62) 4-Bromofluorobenzene	11.061	95	120479	53.026	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.060%
Target Compounds						
						Qvalue
30) Chloroform	5.074	83	1408	0.442	ug/l	# 64
44) Trichloroethene	7.117	130	38283	20.622	ug/l	92
64) Tetrachloroethene	9.263	164	2610	1.532	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093025\
Data File : VX047911.D
Acq On : 30 Sep 2025 15:19
Operator : JC/MD
Sample : Q3193-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GM38-RW3-MW1-20250922

Quant Time: Oct 01 01:12:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

