

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093025\
 Data File : VX047913.D
 Acq On : 30 Sep 2025 16:00
 Operator : JC/MD
 Sample : Q3193-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW10A-MW01I-20250923

Quant Time: Oct 01 01:13:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	125955	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	261894	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	270426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	134429	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	101063	50.408	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.820%
35) Dibromofluoromethane	5.373	113	84978	48.382	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.760%
50) Toluene-d8	8.635	98	294330	48.429	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.860%
62) 4-Bromofluorobenzene	11.061	95	123994	53.758	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.520%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1944	2.230	ug/l	# 80
44) Trichloroethene	7.117	130	683	0.362	ug/l	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093025\
Data File : VX047913.D
Acq On : 30 Sep 2025 16:00
Operator : JC/MD
Sample : Q3193-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW10A-MW01I-20250923

Quant Time: Oct 01 01:13:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

