

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093025\
 Data File : VX047917.D
 Acq On : 30 Sep 2025 17:24
 Operator : JC/MD
 Sample : Q3193-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW10A-MW01S-20250923

Quant Time: Oct 01 01:14:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	123892	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	253037	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	261549	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	128891	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	102125	51.786	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.580%	
35) Dibromofluoromethane	5.379	113	85539	50.406	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.820%	
50) Toluene-d8	8.635	98	290511	49.474	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.940%	
62) 4-Bromofluorobenzene	11.067	95	122391	54.920	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.840%	
Target Compounds						Qvalue
16) Acetone	2.380	43	2079	2.425	ug/l	90
22) Diisopropyl ether	3.751	45	12694	2.156	ug/l #	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093025\
Data File : VX047917.D
Acq On : 30 Sep 2025 17:24
Operator : JC/MD
Sample : Q3193-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW10A-MW01S-20250923

Quant Time: Oct 01 01:14:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

