

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093025\
 Data File : VX047919.D
 Acq On : 30 Sep 2025 18:06
 Operator : JC/MD
 Sample : Q3193-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP11-20250923

Quant Time: Oct 01 01:15:35 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	126991	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.750	114	265107	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.036	117	272232	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	135633	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	105347	52.116	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.240%
35) Dibromofluoromethane	5.379	113	87043	48.957	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.920%
50) Toluene-d8	8.634	98	295151	47.976	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.960%
62) 4-Bromofluorobenzene	11.067	95	124569	53.353	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.700%
Target Compounds						
16) Acetone	2.373	43	1500	1.707	ug/l	Qvalue # 80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX093025\
Data File : VX047919.D
Acq On : 30 Sep 2025 18:06
Operator : JC/MD
Sample : Q3193-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP11-20250923

Quant Time: Oct 01 01:15:35 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

