

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100323\
 Data File : VX038048.D
 Acq On : 03 Oct 2023 19:39
 Operator : JC/MD
 Sample : 04566-05ME
 Misc : 4.55g/5.0 mL/100uL/5.0mL/MSVOA_X/MeOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 B102-08-1(8-8.5)ME

Quant Time: Oct 04 09:15:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 04 09:08:35 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	121076	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	209048	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	192808	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104711	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	98175	52.113	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.220%	
35) Dibromofluoromethane	5.391	113	68645	47.864	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.720%	
50) Toluene-d8	8.647	98	270613	49.305	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 98.620%	
62) 4-Bromofluorobenzene	11.085	95	97309	48.715	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.440%	
Target Compounds						
					Qvalue	
17) Carbon Disulfide	2.483	76	3840	1.184	ug/l	98
18) Methyl Acetate	2.721	43	11166	3.139	ug/l	99
31) Cyclohexane	5.458	56	6382	2.897	ug/l	91
39) Methylcyclohexane	7.373	83	4986	1.996	ug/l #	87
40) Benzene	6.038	78	31034	5.205	ug/l	99
65) Chlorobenzene	10.080	112	6081	1.502	ug/l	100
67) Ethyl Benzene	10.195	91	20300	2.788	ug/l	100
68) m/p-Xylenes	10.305	106	7669	2.750	ug/l	91
69) o-Xylene	10.647	106	1922m	0.726	ug/l	
73) Isopropylbenzene	10.964	105	10587	1.490	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	18519	3.036	ug/l	95
95) Naphthalene	13.780	128	10209	1.313	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100323\
Data File : VX038048.D
Acq On : 03 Oct 2023 19:39
Operator : JC/MD
Sample : 04566-05ME
Misc : 4.55g/5.0 mL/100uL/5.0mL/MSVOA_X/MeOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
B102-08-1(8-8.5)ME

Quant Time: Oct 04 09:15:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 04 09:08:35 2023
Response via : Initial Calibration

