

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100323\
 Data File : VX038037.D
 Acq On : 03 Oct 2023 13:46
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Oct 03 15:16:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 03 13:58:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	120080	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	214111	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197676	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96440	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	87190	44.843	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	89.680%		
35) Dibromofluoromethane	5.391	113	68837	44.830	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	89.660%		
50) Toluene-d8	8.647	98	262964	44.325	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	88.640%#		
62) 4-Bromofluorobenzene	11.079	95	93864	42.739	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	85.480%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.166	85	1037	0.818	ug/l	97
3) Chloromethane	1.300	50	1106	Below Cal	#	83
4) Vinyl Chloride	1.374	62	725	0.488	ug/l #	46
5) Bromomethane	1.611	94	703	4.449	ug/l #	79
6) Chloroethane	1.538	64	6953	6.230	ug/l #	48
7) Trichlorofluoromethane	1.892	101	1531	0.563	ug/l #	83
9) 1,1,2-Trichlorotrifluo...	2.331	101	782	0.651	ug/l #	93
10) Methyl Iodide	2.453	142	616	5.991	ug/l #	86
11) Tert butyl alcohol	2.946	59	1530	3.734	ug/l #	78
12) 1,1-Dichloroethene	2.325	96	825	0.690	ug/l #	68
13) Acrolein	2.233	56	487	2.127	ug/l #	82
14) Allyl chloride	2.666	41	994	0.473	ug/l	93
15) Acrylonitrile	3.074	53	2113	2.333	ug/l	96
16) Acetone	2.379	43	1408	1.713	ug/l #	79
17) Carbon Disulfide	2.514	76	8856	2.636	ug/l	96
18) Methyl Acetate	2.703	43	1031	0.282	ug/l #	85
20) Methylene Chloride	2.788	84	1474	0.945	ug/l	89
21) trans-1,2-Dichloroethene	3.099	96	1722	1.204	ug/l #	83
23) Vinyl Acetate	3.745	43	1856	0.579	ug/l #	73
25) 2-Butanone	4.586	43	1070	0.812	ug/l #	51
27) cis-1,2-Dichloroethene	4.501	96	871	0.523	ug/l	84
28) Bromochloromethane	4.903	49	341	0.273	ug/l #	91
29) Tetrahydrofuran	5.019	42	365	0.422	ug/l #	65
31) Cyclohexane	5.458	56	702	0.318	ug/l #	80
32) 1,1,1-Trichloroethane	5.385	97	538	0.222	ug/l #	45
36) 1,1-Dichloropropene	5.702	75	1809	0.835	ug/l #	72
38) Carbon Tetrachloride	5.684	117	910	0.418	ug/l #	58
39) Methylcyclohexane	7.379	83	1419	0.550	ug/l #	81
40) Benzene	6.050	78	1637	0.262	ug/l #	89
42) 1,2-Dichloroethane	6.098	62	686	0.278	ug/l #	77
44) Trichloroethene	7.135	130	1206	0.747	ug/l	97
46) Dibromomethane	7.586	93	578	0.478	ug/l #	83
49) 1,4-Dioxane	7.689	88	225	5.475	ug/l #	80
51) 4-Methyl-2-Pentanone	8.580	43	917	0.339	ug/l	100
52) Toluene	8.726	92	1620	0.400	ug/l	87
53) t-1,3-Dichloropropene	8.982	75	966	2.309	ug/l	89

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100323\
 Data File : VX038037.D
 Acq On : 03 Oct 2023 13:46
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Oct 03 15:16:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 03 13:58:28 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
54) cis-1,3-Dichloropropene	8.372	75	760	0.303	ug/l #	84
58) 2-Chloroethyl Vinyl ether	8.256	63	691	0.554	ug/l #	55
59) 2-Hexanone	9.439	43	1372	0.638	ug/l	96
61) 1,2-Dibromoethane	9.610	107	479	0.262	ug/l	81
64) Tetrachloroethene	9.275	164	1147	0.874	ug/l	86
65) Chlorobenzene	10.085	112	2497	0.593	ug/l	99
67) Ethyl Benzene	10.195	91	3886	0.515	ug/l	95
68) m/p-Xylenes	10.305	106	2858	0.990	ug/l	87
69) o-Xylene	10.640	106	931	0.336	ug/l	86
70) Styrene	10.658	104	2236	0.485	ug/l #	91
71) Bromoform	10.799	173	211	2.791	ug/l #	29
73) Isopropylbenzene	10.963	105	3230	0.478	ug/l	96
76) 1,2,3-Trichloropropane	11.079	75	52879	25.809	ug/l #	38
77) Bromobenzene	11.201	156	1018	0.622	ug/l	90
78) n-propylbenzene	11.305	91	5922	0.723	ug/l	98
79) 2-Chlorotoluene	11.366	91	3381	0.698	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	3378	0.587	ug/l	90
81) trans-1,4-Dichloro-2-b...	11.079	75	52879	65.050	ug/l #	5
82) 4-Chlorotoluene	11.457	91	5199	0.906	ug/l	97
83) tert-Butylbenzene	11.713	119	2357	0.410	ug/l	91
84) 1,2,4-Trimethylbenzene	11.756	105	3510	0.598	ug/l	95
85) sec-Butylbenzene	11.890	105	4253	0.572	ug/l	96
86) p-Isopropyltoluene	12.012	119	4295	0.698	ug/l	93
87) 1,3-Dichlorobenzene	11.969	146	3297	1.034	ug/l	98
88) 1,4-Dichlorobenzene	11.969	146	3297	1.017	ug/l	97
89) n-Butylbenzene	12.335	91	5645	0.932	ug/l	97
90) Hexachloroethane	12.536	117	416	0.401	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	2323	0.734	ug/l	94
92) 1,2-Dibromo-3-Chloropr...	12.945	75	245	0.418	ug/l	69
93) 1,2,4-Trichlorobenzene	13.591	180	3830	1.821	ug/l	96
94) Hexachlorobutadiene	13.725	225	1913	2.167	ug/l	98
95) Naphthalene	13.774	128	8988	1.233	ug/l	98
96) 1,2,3-Trichlorobenzene	13.963	180	3150	1.494	ug/l	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100323\
 Data File : VX038037.D
 Acq On : 03 Oct 2023 13:46
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Oct 03 15:16:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100323W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 03 13:58:28 2023
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion:168 Resp: 120080
Ion Ratio Lower Upper
168 100
99 60.2 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 0.818 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 85 Resp: 1037
Ion Ratio Lower Upper
85 100
87 32.0 16.7 50.1





#3

Chloromethane

Concen: Below Cal

RT: 1.300 min Scan# 35

Delta R.T. -0.001 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46

Instrument :

MSVOA_X

ClientSampleId :





VIBLK

Tgt Ion: 50 Resp: 1106

Ion Ratio Lower Upper

50 100

52 22.3 25.5 38.3#



#4

Vinyl Chloride

Concen: 0.488 ug/l

RT: 1.374 min Scan# 47

Delta R.T. -0.000 min

Lab File: VX038037.D




Acq: 03 Oct 2023 13:46


Tgt Ion: 62 Resp: 725

Ion Ratio Lower Upper

62 100

64 61.7 25.4 38.0#









#5
Bromomethane
Concen: 4.449 ug/l
RT: 1.611 min Scan# 80
Delta R.T. 0.012 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion: 94 Resp: 703
Ion Ratio Lower Upper
94 100
96 73.2 74.6 111.8#



#6
Chloroethane
Concen: 6.230 ug/l
RT: 1.538 min Scan# 74
Delta R.T. -0.147 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 64 Resp: 6953
Ion Ratio Lower Upper
64 100
66 2.7 25.1 37.7#









#7
 Trichlorofluoromethane
 Concen: 0.563 ug/l
 RT: 1.892 min Scan# 11
 Delta R.T. 0.006 min
 Lab File: VX038037.D
 Acq: 03 Oct 2023 13:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK


Tgt Ion:101 Resp: 1531
 Ion Ratio Lower Upper
 101 100
 103 77.7 51.4 77.2#



#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 0.651 ug/l
 RT: 2.331 min Scan# 204
 Delta R.T. -0.000 min
 Lab File: VX038037.D
 Acq: 03 Oct 2023 13:46


Tgt Ion:101 Resp: 782
 Ion Ratio Lower Upper
 101 100
 85 35.5 37.0 55.6#
 151 72.5 58.6 87.8








#10
Methyl Iodide
Concen: 5.991 ug/l
RT: 2.453 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Instrument :
MSVOA_X
ClientSampleId :
VIBLK







Tgt Ion:142 Resp: 616
Ion Ratio Lower Upper
142 100
127 41.9 41.6 62.4
141 9.7 11.6 17.4#



#11
Tert butyl alcohol
Concen: 3.734 ug/l
RT: 2.946 min Scan# 305
Delta R.T. -0.013 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 59 Resp: 1530
Ion Ratio Lower Upper
59 100
57 0.0 6.0 9.0#





#12

1,1-Dichloroethene

Concen: 0.690 ug/l

RT: 2.325 min Scan# 202

Delta R.T. 0.006 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46

Instrument :

MSVOA_X

ClientSampleId :

VIBLK





Tgt Ion: 96 Resp: 825

Ion Ratio Lower Upper

96 100

61 139.5 144.6 216.8#

98 94.2 50.4 75.6#



#13

Acrolein

Concen: 2.127 ug/l

RT: 2.233 min Scan# 188

Delta R.T. 0.000 min

Lab File: VX038037.D




Acq: 03 Oct 2023 13:46


Tgt Ion: 56 Resp: 487

Ion Ratio Lower Upper

56 100

55 86.2 57.1 85.7#









#14
 Allyl chloride
 Concen: 0.473 ug/l
 RT: 2.666 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VX038037.D
 Acq: 03 Oct 2023 13:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK


Tgt Ion:	41	Resp:	994
Ion	Ratio	Lower	Upper
41	100		
39	77.1	55.8	83.6
76	28.9	24.3	36.5



#15
 Acrylonitrile
 Concen: 2.333 ug/l
 RT: 3.074 min Scan# 326
 Delta R.T. 0.012 min
 Lab File: VX038037.D
 Acq: 03 Oct 2023 13:46

Tgt Ion:	53	Resp:	2113
Ion	Ratio	Lower	Upper
53	100		
52	79.6	66.0	99.0
51	40.0	29.8	44.6









#16
Acetone
Concen: 1.713 ug/l
RT: 2.379 min Scan# 212
Delta R.T. -0.001 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion: 43 Resp: 1408
Ion Ratio Lower Upper
43 100
58 38.5 21.9 32.9#



#17
Carbon Disulfide
Concen: 2.636 ug/l
RT: 2.514 min Scan# 234
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 76 Resp: 8856
Ion Ratio Lower Upper
76 100
78 10.1 7.0 10.6









#18
Methyl Acetate
Concen: 0.282 ug/l
RT: 2.703 min Scan# 21
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion: 43 Resp: 1031
Ion Ratio Lower Upper
43 100
74 29.7 17.8 26.8#



#20
Methylene Chloride
Concen: 0.945 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 84 Resp: 1474
Ion Ratio Lower Upper
84 100
49 110.9 103.6 155.4
51 45.4 33.8 50.6
86 57.9 49.4 74.0









#21
trans-1,2-Dichloroethene
Concen: 1.204 ug/l
RT: 3.099 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion: 96 Resp: 1722
Ion Ratio Lower Upper
96 100
61 168.4 119.2 178.8
98 75.9 48.7 73.1#



#23
Vinyl Acetate
Concen: 0.579 ug/l
RT: 3.745 min Scan# 436
Delta R.T. 0.024 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 43 Resp: 1856
Ion Ratio Lower Upper
43 100
86 0.0 7.8 11.8#









#25
2-Butanone
Concen: 0.812 ug/l
RT: 4.586 min Scan# 51
Delta R.T. 0.030 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion: 43 Resp: 1070
Ion Ratio Lower Upper
43 100
72 0.0 19.4 29.2#



#27
cis-1,2-Dichloroethene
Concen: 0.523 ug/l
RT: 4.501 min Scan# 560
Delta R.T. 0.012 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 96 Resp: 871
Ion Ratio Lower Upper
96 100
61 128.0 0.0 312.2
98 65.8 0.0 129.0









#28
Bromochloromethane
Concen: 0.273 ug/l
RT: 4.903 min Scan# 61
Delta R.T. 0.005 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion: 49 Resp: 341
Ion Ratio Lower Upper
49 100
129 0.0 0.0 3.8
130 79.5 57.5 86.3



#29
Tetrahydrofuran
Concen: 0.422 ug/l
RT: 5.019 min Scan# 645
Delta R.T. 0.012 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 42 Resp: 365
Ion Ratio Lower Upper
42 100
72 38.1 34.2 51.4
71 0.0 32.2 48.2#









#31
Cyclohexane
Concen: 0.318 ug/l
RT: 5.458 min Scan# 719
Delta R.T. -0.013 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion: 56 Resp: 702
Ion Ratio Lower Upper
56 100
69 36.9 23.4 35.2#
84 62.8 65.9 98.9#



#32
1,1,1-Trichloroethane
Concen: 0.222 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 97 Resp: 538
Ion Ratio Lower Upper
97 100
99 0.0 51.1 76.7#
61 34.0 36.4 54.6#









#33
1,2-Dichloroethane-d4
Concen: 44.843 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion: 65 Resp: 87190
Ion Ratio Lower Upper
65 100
67 51.6 0.0 100.2



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 931
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 114 Resp: 214111
Ion Ratio Lower Upper
114 100
63 21.1 0.0 45.0
88 15.2 0.0 36.6









#35
Dibromofluoromethane
Concen: 44.830 ug/l
RT: 5.391 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion:113 Resp: 68837
Ion Ratio Lower Upper
113 100
111 102.2 82.9 124.3
192 17.5 13.3 19.9



#36
1,1-Dichloropropene
Concen: 0.835 ug/l
RT: 5.702 min Scan# 757
Delta R.T. 0.006 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 75 Resp: 1809
Ion Ratio Lower Upper
75 100
110 18.5 17.0 50.9
77 14.5 24.7 37.1#









#38
Carbon Tetrachloride
Concen: 0.418 ug/l
RT: 5.684 min Scan# 753
Delta R.T. 0.006 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion:117 Resp: 910
Ion Ratio Lower Upper
117 100
119 60.1 78.2 117.2#
121 0.0 24.4 36.6#



#39
Methylcyclohexane
Concen: 0.550 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 83 Resp: 1419
Ion Ratio Lower Upper
83 100
55 62.9 64.6 97.0#
98 35.2 36.7 55.1#





#40

Benzene

Concen: 0.262 ug/l

RT: 6.050 min Scan# 812

Delta R.T. 0.012 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46

Instrument :

MSVOA_X

ClientSampleId :





VIBLK

Tgt Ion: 78 Resp: 1637

Ion Ratio Lower Upper

78 100

77 18.5 19.3 28.9#



#42

1,2-Dichloroethane

Concen: 0.278 ug/l

RT: 6.098 min Scan# 822

Delta R.T. 0.012 min

Lab File: VX038037.D




Acq: 03 Oct 2023 13:46


Tgt Ion: 62 Resp: 686

Ion Ratio Lower Upper

62 100

98 0.0 0.0 16.0









#44
Trichloroethene
Concen: 0.747 ug/l
RT: 7.135 min Scan# 991
Delta R.T. 0.006 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion:130 Resp: 1206
Ion Ratio Lower Upper
130 100
95 98.9 0.0 204.4



#46
Dibromomethane
Concen: 0.478 ug/l
RT: 7.586 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 93 Resp: 578
Ion Ratio Lower Upper
93 100
95 75.3 66.6 100.0
174 64.0 69.4 104.0#





#49

1,4-Dioxane

Concen: 5.475 ug/l

RT: 7.689 min Scan# 1078

Delta R.T. 0.030 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46

Instrument :

MSVOA_X

ClientSampleId :

VIBLK





Tgt Ion: 88 Resp: 225

Ion Ratio Lower Upper

88 100

43 0.0 0.0 0.0

58 57.8 60.1 90.1#



#50

Toluene-d8

Concen: 44.325 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.006 min

Lab File: VX038037.D




Acq: 03 Oct 2023 13:46


Tgt Ion: 98 Resp: 262964

Ion Ratio Lower Upper

98 100

100 63.2 53.1 79.7









#51
4-Methyl-2-Pentanone
Concen: 0.339 ug/l
RT: 8.580 min Scan# 1229
Delta R.T. 0.006 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion: 43 Resp: 917
Ion Ratio Lower Upper
43 100
58 38.1 30.4 45.6



#52
Toluene
Concen: 0.400 ug/l
RT: 8.726 min Scan# 1253
Delta R.T. 0.006 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 92 Resp: 1620
Ion Ratio Lower Upper
92 100
91 185.6 134.2 201.2





#53

t-1,3-Dichloropropene

Concen: 2.309 ug/l

RT: 8.982 min Scan# 1195

Delta R.T. -0.000 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46

Instrument :

MSVOA_X

ClientSampleId :





VIBLK

Tgt Ion: 75 Resp: 966

Ion Ratio Lower Upper

75 100

77 38.0 25.7 38.5



#54

cis-1,3-Dichloropropene

Concen: 0.303 ug/l

RT: 8.372 min Scan# 1195

Delta R.T. 0.006 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46




Tgt Ion: 75 Resp: 760


Ion Ratio Lower Upper

75 100

77 19.3 25.3 37.9#

39 68.3 46.7 70.1









#58
2-Chloroethyl Vinyl ether
Concen: 0.554 ug/l
RT: 8.256 min Scan# 1176
Delta R.T. 0.012 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion: 63 Resp: 691
Ion Ratio Lower Upper
63 100
106 4.2 22.5 33.7#



#59
2-Hexanone
Concen: 0.638 ug/l
RT: 9.439 min Scan# 1370
Delta R.T. 0.006 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 43 Resp: 1372
Ion Ratio Lower Upper
43 100
58 49.4 26.2 78.5





#61

1,2-Dibromoethane

Concen: 0.262 ug/l

RT: 9.610 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46

Instrument :

MSVOA_X

ClientSampleId :





VIBLK

Tgt Ion:107 Resp: 479

Ion Ratio Lower Upper

107 100

109 75.4 75.2 112.8



#62

4-Bromofluorobenzene

Concen: 42.739 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. 0.000 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46




Tgt Ion: 95 Resp: 93864


Ion Ratio Lower Upper

95 100

174 78.6 0.0 135.6

176 76.1 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion:117 Resp: 197676
Ion Ratio Lower Upper
117 100
82 58.0 46.2 69.4
119 32.1 25.5 38.3



#64
Tetrachloroethene
Concen: 0.874 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion:164 Resp: 1147
Ion Ratio Lower Upper
164 100
166 110.7 102.7 154.1
129 92.9 80.4 120.6
131 78.6 76.5 114.7





#65

Chlorobenzene

Concen: 0.593 ug/l

RT: 10.085 min Scan# 1475

Delta R.T. 0.005 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46

Instrument :

MSVOA_X

ClientSampleId :





VIBLK

Tgt Ion:112 Resp: 2497

Ion Ratio Lower Upper

112 100

114 31.3 25.7 38.5



#67

Ethyl Benzene

Concen: 0.515 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. 0.000 min

Lab File: VX038037.D




Acq: 03 Oct 2023 13:46


Tgt Ion: 91 Resp: 3886

Ion Ratio Lower Upper

91 100

106 33.9 24.9 37.3









#68
m/p-Xylenes
Concen: 0.990 ug/l
RT: 10.305 min Scan# 1511
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion:106 Resp: 2858
Ion Ratio Lower Upper
106 100
91 227.7 165.5 248.3



#69
o-Xylene
Concen: 0.336 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion:106 Resp: 931
Ion Ratio Lower Upper
106 100
91 241.0 108.9 326.8









#70
Styrene
Concen: 0.485 ug/l
RT: 10.658 min Scan# 1569
Delta R.T. 0.005 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion:104 Resp: 2236
Ion Ratio Lower Upper
104 100
78 43.3 43.6 65.4#
103 53.0 43.6 65.4



#71
Bromoform
Concen: 2.791 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion:173 Resp: 211
Ion Ratio Lower Upper
173 100
175 0.0 24.1 72.4#
254 0.0 0.0 0.0





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1

Delta R.T. 0.000 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46

Instrument :

MSVOA_X

ClientSampleId :

VIBLK





Tgt Ion:152 Resp: 96440

Ion Ratio Lower Upper

152 100

115 61.8 43.3 129.9

150 158.0 0.0 345.0



#73

Isopropylbenzene

Concen: 0.478 ug/l

RT: 10.963 min Scan# 1620

Delta R.T. -0.001 min

Lab File: VX038037.D




Acq: 03 Oct 2023 13:46


Tgt Ion:105 Resp: 3230

Ion Ratio Lower Upper

105 100

120 27.6 12.9 38.6





#76

1,2,3-Trichloropropane

Concen: 25.809 ug/l

RT: 11.079 min Scan# 1639

Delta R.T. -0.159 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46

Instrument :

MSVOA_X

ClientSampleId :





VIBLK

Tgt Ion: 75 Resp: 52879

Ion Ratio Lower Upper

75 100

77 1.0 19.1 57.5#



#77

Bromobenzene

Concen: 0.622 ug/l

RT: 11.201 min Scan# 1659

Delta R.T. 0.000 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46




Tgt Ion: 156 Resp: 1018


Ion Ratio Lower Upper

156 100

77 156.9 88.4 265.3

158 98.3 48.0 144.0









#78
n-propylbenzene
Concen: 0.723 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion: 91 Resp: 5922
Ion Ratio Lower Upper
91 100
120 21.7 11.5 34.4



#79
2-Chlorotoluene
Concen: 0.698 ug/l
RT: 11.366 min Scan# 1686
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 91 Resp: 3381
Ion Ratio Lower Upper
91 100
126 32.2 16.3 48.8









#80
 1,3,5-Trimethylbenzene
 Concen: 0.587 ug/l
 RT: 11.451 min Scan# 11
 Delta R.T. 0.000 min
 Lab File: VX038037.D
 Acq: 03 Oct 2023 13:46




Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK


Tgt Ion:105 Resp: 3378
 Ion Ratio Lower Upper
 105 100
 120 41.6 24.1 72.3



#81
 trans-1,4-Dichloro-2-butene
 Concen: 65.050 ug/l
 RT: 11.079 min Scan# 1639
 Delta R.T. 0.061 min
 Lab File: VX038037.D
 Acq: 03 Oct 2023 13:46

Tgt Ion: 75 Resp: 52879
 Ion Ratio Lower Upper
 75 100
 53 0.0 90.5 135.7#
 89 0.0 36.9 55.3#





#82

4-Chlorotoluene

Concen: 0.906 ug/l

RT: 11.457 min Scan# 1700

Delta R.T. 0.000 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46

Instrument :

MSVOA_X

ClientSampleId :





VIBLK

Tgt Ion: 91 Resp: 5199

Ion Ratio Lower Upper

91 100

126 26.9 14.2 42.6



#83

tert-Butylbenzene

Concen: 0.410 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. 0.000 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46




Tgt Ion: 119 Resp: 2357


Ion Ratio Lower Upper

119 100

91 70.1 31.2 93.6

134 26.7 11.7 35.0









#84
1,2,4-Trimethylbenzene
Concen: 0.598 ug/l
RT: 11.756 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion:105 Resp: 3510
Ion Ratio Lower Upper
105 100
120 45.8 21.2 63.6



#85
sec-Butylbenzene
Concen: 0.572 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion:105 Resp: 4253
Ion Ratio Lower Upper
105 100
134 17.9 9.8 29.4









#86
p-Isopropyltoluene
Concen: 0.698 ug/l
RT: 12.012 min Scan# 1792
Delta R.T. -0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion:119	Resp:	4295
Ion	Ratio	Lower Upper
119	100	
134	24.3	12.8 38.3
91	32.1	13.1 39.3



#87
1,3-Dichlorobenzene
Concen: 1.034 ug/l
RT: 11.969 min Scan# 1785
Delta R.T. 0.000 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion:146	Resp:	3297
Ion	Ratio	Lower Upper
146	100	
111	45.1	22.1 66.1
148	64.9	31.8 95.3





#88

1,4-Dichlorobenzene

Concen: 1.017 ug/l

RT: 11.969 min Scan# 11

Delta R.T. -0.074 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46

Instrument :

MSVOA_X

ClientSampleId :

VIBLK





Tgt Ion:146 Resp: 3297

Ion Ratio Lower Upper

146 100

111 45.1 21.3 63.7

148 64.9 31.8 95.4



#89

n-Butylbenzene

Concen: 0.932 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46




Tgt Ion: 91 Resp: 5645


Ion Ratio Lower Upper

91 100

92 53.1 27.4 82.0

134 22.2 12.6 37.8





#90

Hexachloroethane

Concen: 0.401 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. 0.000 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46

Instrument :

MSVOA_X

ClientSampleId :





VIBLK

Tgt Ion:117 Resp: 416

Ion Ratio Lower Upper

117 100

201 66.3 32.9 98.6



#91

1,2-Dichlorobenzene

Concen: 0.734 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.000 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46




Tgt Ion:146 Resp: 2323


Ion Ratio Lower Upper

146 100

111 39.9 22.4 67.0

148 59.7 31.6 95.0









#92
1,2-Dibromo-3-Chloropropane
Concen: 0.418 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. 0.006 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46




Instrument :
MSVOA_X
ClientSampleId :
VIBLK


Tgt Ion: 75 Resp: 245
Ion Ratio Lower Upper
75 100
155 41.2 36.1 108.5
157 65.3 44.9 134.7



#93
1,2,4-Trichlorobenzene
Concen: 1.821 ug/l
RT: 13.591 min Scan# 2051
Delta R.T. 0.006 min
Lab File: VX038037.D
Acq: 03 Oct 2023 13:46

Tgt Ion: 180 Resp: 3830
Ion Ratio Lower Upper
180 100
182 94.4 48.3 144.8
145 30.2 18.0 54.0





#94

Hexachlorobutadiene

Concen: 2.167 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.000 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46

Instrument :

MSVOA_X

ClientSampleId :

VIBLK





Tgt Ion:225 Resp: 1913

Ion Ratio Lower Upper

225 100

223 64.4 31.1 93.2

227 64.0 32.5 97.5



#95

Naphthalene

Concen: 1.233 ug/l

RT: 13.774 min Scan# 2081

Delta R.T. -0.000 min

Lab File: VX038037.D

Acq: 03 Oct 2023 13:46




Tgt Ion:128 Resp: 8988


Ion Ratio Lower Upper

128 100

127 11.8 10.3 15.5


129 10.3 8.8 13.2





#96
 1,2,3-Trichlorobenzene
 Concen: 1.494 ug/l
 RT: 13.963 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: VX038037.D
 Acq: 03 Oct 2023 13:46

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK



Tgt Ion:	180	Resp:	3150
Ion	Ratio	Lower	Upper
180	100		
182	103.9	46.9	140.6
145	35.3	17.8	53.5

