

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101323\
 Data File : VX038285.D
 Acq On : 13 Oct 2023 15:29
 Operator : JC/MD
 Sample : 04889-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 13 23:13:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	105074	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	190328	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199109	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103013	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	78741	46.591	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.180%
35) Dibromofluoromethane	5.385	113	62709	45.725	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.460%
50) Toluene-d8	8.647	98	244670	46.688	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.380%
62) 4-Bromofluorobenzene	11.079	95	93940	47.283	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.560%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101323\
Data File : VX038285.D
Acq On : 13 Oct 2023 15:29
Operator : JC/MD
Sample : 04889-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Oct 13 23:13:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

