

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101323\
 Data File : VX038288.D
 Acq On : 13 Oct 2023 16:39
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Oct 13 23:14:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	108480	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	195037	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206857	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108966	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	80136	45.928	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.860%
35) Dibromofluoromethane	5.385	113	64372	45.805	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.600%
50) Toluene-d8	8.647	98	251742	46.877	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.760%
62) 4-Bromofluorobenzene	11.079	95	97794	48.034	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101323\
Data File : VX038288.D
Acq On : 13 Oct 2023 16:39
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Oct 13 23:14:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

