

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101323\
 Data File : VX038284.D
 Acq On : 13 Oct 2023 15:06
 Operator : JC/MD
 Sample : 04889-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 13 23:12:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	111336	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	201923	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208344	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108260	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	82813	46.244	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.480%
35) Dibromofluoromethane	5.385	113	66942	46.009	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.020%
50) Toluene-d8	8.647	98	259047	46.592	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.180%
62) 4-Bromofluorobenzene	11.079	95	98349	46.660	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.320%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101323\
Data File : VX038284.D
Acq On : 13 Oct 2023 15:06
Operator : JC/MD
Sample : 04889-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Oct 13 23:12:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 711
Delta R.T. -0.000 min
Lab File: VX038284.D
Acq: 13 Oct 2023 15:06




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5


Tgt Ion:168 Resp: 111336
Ion Ratio Lower Upper
168 100
99 58.0 56.6 85.0



#33
1,2-Dichloroethane-d4
Concen: 46.244 ug/l
RT: 5.958 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX038284.D
Acq: 13 Oct 2023 15:06

Tgt Ion: 65 Resp: 82813
Ion Ratio Lower Upper
65 100
67 51.5 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 911
Delta R.T. -0.000 min
Lab File: VX038284.D
Acq: 13 Oct 2023 15:06




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5


Tgt Ion:114 Resp: 201923
Ion Ratio Lower Upper
114 100
63 19.7 0.0 45.0
88 15.8 0.0 36.6



#35
Dibromofluoromethane
Concen: 46.009 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX038284.D
Acq: 13 Oct 2023 15:06

Tgt Ion:113 Resp: 66942
Ion Ratio Lower Upper
113 100
111 102.3 82.9 124.3
192 17.7 13.3 19.9









#50
Toluene-d8
Concen: 46.592 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX038284.D
Acq: 13 Oct 2023 15:06




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5


Tgt Ion: 98 Resp: 259047
Ion Ratio Lower Upper
98 100
100 62.9 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 46.660 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX038284.D
Acq: 13 Oct 2023 15:06

Tgt Ion: 95 Resp: 98349
Ion Ratio Lower Upper
95 100
174 80.4 0.0 135.6
176 76.9 0.0 131.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. -0.000 min

Lab File: VX038284.D

Acq: 13 Oct 2023 15:06

Instrument :

MSVOA_X

ClientSampleId :

STORAGE-BLANK-WATER-REF-5





Tgt Ion:117 Resp: 208344

Ion Ratio Lower Upper

117 100

82 54.5 46.2 69.4

119 31.6 25.5 38.3



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX038284.D

Acq: 13 Oct 2023 15:06

Tgt Ion:152 Resp: 108260

Ion Ratio Lower Upper

152 100

115 60.3 43.3 129.9

150 158.7 0.0 345.0

