



BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101323\
Data File : VX038271.D
Acq On : 13 Oct 2023 08:31
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Title : SW846 8260
Last Update : Thu Oct 05 15:27:13 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.3	8517	PASS
75	95	30	60	55.2	22040	PASS
95	95	100	100	100.0	39907	PASS
96	95	5	9	6.9	2773	PASS
173	174	0.00	2	1.2	376	PASS
174	95	50	100	79.3	31645	PASS
175	174	5	9	7.6	2400	PASS
176	174	95	101	95.0	30075	PASS
177	176	5	9	6.5	1950	PASS