

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
 Data File : VX038295.D
 Acq On : 16 Oct 2023 11:35
 Operator : JC/MD
 Sample : VX1016MBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1016MBL01

Quant Time: Oct 17 01:18:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	150127	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	268427	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	272498	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	141730	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	104119	43.119	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.240%
35) Dibromofluoromethane	5.385	113	85317	44.110	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.220%
50) Toluene-d8	8.647	98	337027	45.600	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.200%#
62) 4-Bromofluorobenzene	11.079	95	133496	47.643	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
Data File : VX038295.D
Acq On : 16 Oct 2023 11:35
Operator : JC/MD
Sample : VX1016MBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1016MBL01

Quant Time: Oct 17 01:18:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

