

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
 Data File : VX038301.D
 Acq On : 16 Oct 2023 14:11
 Operator : JC/MD
 Sample : 04899-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Oct 17 01:22:19 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	131129	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	237459	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	246373	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129512	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95479	45.269	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.540%
35) Dibromofluoromethane	5.385	113	75734	44.262	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.520%
50) Toluene-d8	8.647	98	300408	45.946	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.900%#
62) 4-Bromofluorobenzene	11.079	95	118708	47.890	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.780%
Target Compounds						
16) Acetone	2.380	43	2036	2.313	ug/l	Qvalue # 69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
Data File : VX038301.D
Acq On : 16 Oct 2023 14:11
Operator : JC/MD
Sample : 04899-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Oct 17 01:22:19 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

