

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
 Data File : VX038302.D
 Acq On : 16 Oct 2023 14:34
 Operator : JC/MD
 Sample : 04899-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 FB-101323

Quant Time: Oct 17 01:22:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.550	168	132841	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	238192	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	255519	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138516	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96942	45.371	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.740%
35) Dibromofluoromethane	5.385	113	79088	46.080	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.160%
50) Toluene-d8	8.647	98	299387	45.649	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.300%#
62) 4-Bromofluorobenzene	11.079	95	126701	50.958	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.920%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
 Data File : VX038302.D
 Acq On : 16 Oct 2023 14:34
 Operator : JC/MD
 Sample : 04899-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-101323

Quant Time: Oct 17 01:22:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

