

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
 Data File : VX038305.D
 Acq On : 16 Oct 2023 15:46
 Operator : JC/MD
 Sample : 04899-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-15-37-40

Quant Time: Oct 17 01:23:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	167484	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	290639	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	287417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	146693	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	124839	46.342	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.680%
35) Dibromofluoromethane	5.385	113	97315	46.468	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.940%
50) Toluene-d8	8.647	98	353938	44.228	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	88.460%#
62) 4-Bromofluorobenzene	11.079	95	135451	44.646	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.300%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.495	96	4474	1.980	ug/l	75
44) Trichloroethene	7.129	130	79763	37.755	ug/l	100
64) Tetrachloroethene	9.275	164	11371	6.111	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
Data File : VX038305.D
Acq On : 16 Oct 2023 15:46
Operator : JC/MD
Sample : 04899-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-15-37-40

Quant Time: Oct 17 01:23:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

