

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
 Data File : VX038308.D
 Acq On : 16 Oct 2023 16:55
 Operator : JC/MD
 Sample : 04899-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-16-37-40

Quant Time: Oct 17 01:25:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	163409	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	289770	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	288868	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	153300	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	138703	52.772	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.540%	
35) Dibromofluoromethane	5.385	113	106400	50.958	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.920%	
50) Toluene-d8	8.647	98	397676	49.842	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.680%	
62) 4-Bromofluorobenzene	11.079	95	153021	50.589	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.180%	
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.495	96	6011	2.727	ug/l	82
44) Trichloroethene	7.129	130	118236	56.134	ug/l	98
64) Tetrachloroethene	9.275	164	20302	10.856	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
Data File : VX038308.D
Acq On : 16 Oct 2023 16:55
Operator : JC/MD
Sample : 04899-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-16-37-40

Quant Time: Oct 17 01:25:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

