

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
 Data File : VX038309.D
 Acq On : 16 Oct 2023 17:18
 Operator : JC/MD
 Sample : 04899-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-17-10-13

Quant Time: Oct 17 01:25:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	144450	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	251728	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	249314	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131701	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	123561	53.181	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 106.360%	
35) Dibromofluoromethane	5.385	113	97001	53.478	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.960%	
50) Toluene-d8	8.647	98	359082	51.807	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.620%	
62) 4-Bromofluorobenzene	11.079	95	138012	52.522	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.040%	
Target Compounds						
					Qvalue	
30) Chloroform	5.093	83	155502	47.295	ug/l	97
47) Bromodichloromethane	7.824	83	4774	1.911	ug/l	99
52) Toluene	8.720	92	59708	13.102	ug/l	95
64) Tetrachloroethene	9.275	164	3140	1.945	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
Data File : VX038309.D
Acq On : 16 Oct 2023 17:18
Operator : JC/MD
Sample : 04899-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-17-10-13

Quant Time: Oct 17 01:25:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

