

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
 Data File : VX038310.D
 Acq On : 16 Oct 2023 17:41
 Operator : JC/MD
 Sample : 04899-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-17-23-26

Quant Time: Oct 17 01:26:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	146049	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	256666	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	258811	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132839	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	109438	46.587	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	93.180%		
35) Dibromofluoromethane	5.385	113	86897	46.986	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.980%		
50) Toluene-d8	8.647	98	320252	45.315	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	90.640%#		
62) 4-Bromofluorobenzene	11.079	95	121934	45.511	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.020%		
Target Compounds					Qvalue	
30) Chloroform	5.092	83	149767	45.052	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
Data File : VX038310.D
Acq On : 16 Oct 2023 17:41
Operator : JC/MD
Sample : 04899-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-17-23-26

Quant Time: Oct 17 01:26:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

