

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
 Data File : VX038311.D
 Acq On : 16 Oct 2023 18:04
 Operator : JC/MD
 Sample : 04899-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-17-37-40

Quant Time: Oct 17 01:26:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	146105	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	262102	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	259887	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134777	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	126849	53.978	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.960%	
35) Dibromofluoromethane	5.385	113	100171	53.040	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.080%	
50) Toluene-d8	8.647	98	372430	51.606	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.220%	
62) 4-Bromofluorobenzene	11.079	95	138994	50.802	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.600%	
Target Compounds						
						Qvalue
16) Acetone	2.373	43	2481	2.530	ug/l	99
30) Chloroform	5.092	83	60326	18.140	ug/l	100
64) Tetrachloroethene	9.275	164	2066	1.228	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
Data File : VX038311.D
Acq On : 16 Oct 2023 18:04
Operator : JC/MD
Sample : 04899-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-17-37-40

Quant Time: Oct 17 01:26:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

