

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
 Data File : VX038312.D
 Acq On : 16 Oct 2023 18:27
 Operator : JC/MD
 Sample : 04899-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-18-10-13

Quant Time: Oct 17 01:26:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	145802	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	256539	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	254019	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132184	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	125773	53.631	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.260%	
35) Dibromofluoromethane	5.385	113	96850	52.393	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.780%	
50) Toluene-d8	8.647	98	369072	52.249	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.500%	
62) 4-Bromofluorobenzene	11.079	95	138554	51.740	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.480%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	1973	2.016	ug/l	# 88
30) Chloroform	5.093	83	5455	1.644	ug/l	95
44) Trichloroethene	7.129	130	4889	2.622	ug/l	93
52) Toluene	8.720	92	5639	1.214	ug/l	84
64) Tetrachloroethene	9.275	164	11107	6.754	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
Data File : VX038312.D
Acq On : 16 Oct 2023 18:27
Operator : JC/MD
Sample : 04899-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-18-10-13

Quant Time: Oct 17 01:26:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

