

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
 Data File : VX038313.D
 Acq On : 16 Oct 2023 18:50
 Operator : JC/MD
 Sample : 04899-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GW-18-23-26

Quant Time: Oct 17 01:27:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	140095	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	241631	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	243013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	126469	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	118327	52.512	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.020%	
35) Dibromofluoromethane	5.385	113	91696	52.665	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.340%	
50) Toluene-d8	8.647	98	346730	52.115	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.220%	
62) 4-Bromofluorobenzene	11.079	95	129675	51.412	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.820%	
Target Compounds						
						Qvalue
30) Chloroform	5.098	83	75089	23.548	ug/l	97
44) Trichloroethene	7.135	130	1245	0.709	ug/l	74
64) Tetrachloroethene	9.275	164	4079	2.593	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
Data File : VX038313.D
Acq On : 16 Oct 2023 18:50
Operator : JC/MD
Sample : 04899-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GW-18-23-26

Quant Time: Oct 17 01:27:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

