

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
 Data File : VX038302.D
 Acq On : 16 Oct 2023 14:34
 Operator : JC/MD
 Sample : 04899-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-101323

Quant Time: Oct 17 01:22:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	132841	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	238192	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	255519	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	138516	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	96942	45.371	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.740%
35) Dibromofluoromethane	5.385	113	79088	46.080	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.160%
50) Toluene-d8	8.647	98	299387	45.649	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.300%#
62) 4-Bromofluorobenzene	11.079	95	126701	50.958	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.920%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101623\
Data File : VX038302.D
Acq On : 16 Oct 2023 14:34
Operator : JC/MD
Sample : 04899-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-101323

Quant Time: Oct 17 01:22:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.006 min
Lab File: VX038302.D
Acq: 16 Oct 2023 14:34




Instrument :
MSVOA_X
ClientSampleId :
FB-101323


Tgt Ion:168 Resp: 132841
Ion Ratio Lower Upper
168 100
99 57.2 56.6 85.0



#33
1,2-Dichloroethane-d4
Concen: 45.371 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.006 min
Lab File: VX038302.D
Acq: 16 Oct 2023 14:34

Tgt Ion: 65 Resp: 96942
Ion Ratio Lower Upper
65 100
67 51.0 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX038302.D
Acq: 16 Oct 2023 14:34




Instrument :
MSVOA_X
ClientSampleId :
FB-101323


Tgt Ion:114 Resp: 238192
Ion Ratio Lower Upper
114 100
63 20.3 0.0 45.0
88 16.4 0.0 36.6



#35
Dibromofluoromethane
Concen: 46.080 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX038302.D
Acq: 16 Oct 2023 14:34

Tgt Ion:113 Resp: 79088
Ion Ratio Lower Upper
113 100
111 103.0 82.9 124.3
192 17.8 13.3 19.9









#50
Toluene-d8
Concen: 45.649 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX038302.D
Acq: 16 Oct 2023 14:34




Instrument :
MSVOA_X
ClientSampleId :
FB-101323


Tgt Ion: 98 Resp: 299387
Ion Ratio Lower Upper
98 100
100 63.4 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 50.958 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX038302.D
Acq: 16 Oct 2023 14:34

Tgt Ion: 95 Resp: 126701
Ion Ratio Lower Upper
95 100
174 79.6 0.0 135.6
176 76.0 0.0 131.0





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.055 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VX038302.D

Acq: 16 Oct 2023 14:34

Instrument :

MSVOA_X

ClientSampleId :

FB-101323





Tgt Ion:117 Resp: 255519

Ion Ratio Lower Upper

117 100

82 53.6 46.2 69.4

119 31.3 25.5 38.3



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 1794

Delta R.T. 0.000 min

Lab File: VX038302.D

Acq: 16 Oct 2023 14:34

Tgt Ion:152 Resp: 138516

Ion Ratio Lower Upper

152 100

115 60.3 43.3 129.9

150 156.8 0.0 345.0

