



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102324\
Data File : VX043478.D
Acq On : 23 Oct 2024 08:00
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102324W.M
Title : SW846 8260
Last Update : Thu Oct 24 04:54:33 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.9	4299	PASS
75	95	30	60	52.8	11991	PASS
95	95	100	100	100.0	22717	PASS
96	95	5	9	6.7	1517	PASS
173	174	0.00	2	0.7	106	PASS
174	95	50	100	68.7	15601	PASS
175	174	5	9	8.2	1276	PASS
176	174	95	101	99.4	15515	PASS
177	176	5	9	6.0	930	PASS