

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102524\
 Data File : VX043517.D
 Acq On : 25 Oct 2024 12:33
 Operator : JC/MD
 Sample : VX1025MBL02
 Misc : 5.00g/5mL/100uL/5mL/MSVOA_X/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK714

Quant Time: Oct 26 01:35:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML102524WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 26 01:33:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	292279	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	261424	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	121864	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	73462	46.204	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	92.400%
7) Chloroethane-d5	1.666	69	61830	48.559	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.120%
11) 1,1-Dichloroethene-d2	2.306	65	36771	48.974	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	97.940%
21) 2-Butanone-d5	4.458	46	135375	102.215	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	102.220%
24) Chloroform-d	5.056	84	173729	47.243	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.480%
26) 1,2-Dichloroethane-d4	5.952	65	121514	49.179	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.360%
32) Benzene-d6	5.970	84	329839	49.703	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	99.400%
36) 1,2-Dichloropropane-d6	7.306	67	103723	50.210	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	100.420%
41) Toluene-d8	8.647	98	296974	48.879	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.760%
43) trans-1,3-Dichloroprop...	8.952	79	50779	46.723	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.440%
47) 2-Hexanone-d5	9.384	63	109816	107.183	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	107.180%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	155715	50.633	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	101.260%
66) 1,2-Dichlorobenzene-d4	12.317	152	111955	50.615	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102524\
Data File : VX043517.D
Acq On : 25 Oct 2024 12:33
Operator : JC/MD
Sample : VX1025MBL02
Misc : 5.00g/5mL/100ul/5mL/MSVOA_X/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK714

Quant Time: Oct 26 01:35:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM102524WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 26 01:33:53 2024
Response via : Initial Calibration

