

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102524\
 Data File : VX043526.D
 Acq On : 25 Oct 2024 16:24
 Operator : JC/MD
 Sample : P4415-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK002

Quant Time: Oct 26 01:37:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML102524WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 26 01:33:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	247215	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	213715	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	88749	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	77542	57.660	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	115.320%
7) Chloroethane-d5	1.642	69	34645	32.169	ug/L	-0.02
Spiked Amount	50.000	Range	70 - 130	Recovery	=	64.340%#
11) 1,1-Dichloroethene-d2	2.288	65	34530	54.373	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	108.740%
21) 2-Butanone-d5	4.471	46	117115	104.547	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	104.550%
24) Chloroform-d	5.050	84	159523	51.288	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.580%
26) 1,2-Dichloroethane-d4	5.952	65	112862	54.004	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	108.000%
32) Benzene-d6	5.970	84	324474	59.809	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	119.620%
36) 1,2-Dichloropropane-d6	7.306	67	98399	58.266	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	116.540%
41) Toluene-d8	8.647	98	286629	57.708	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	115.420%
43) trans-1,3-Dichloroprop...	8.952	79	40178	45.222	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.440%
47) 2-Hexanone-d5	9.391	63	91908	109.729	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	109.730%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	139877	55.637	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	111.280%
66) 1,2-Dichlorobenzene-d4	12.317	152	97295	60.400	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	120.800%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102524\
Data File : VX043526.D
Acq On : 25 Oct 2024 16:24
Operator : JC/MD
Sample : P4415-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK002

Quant Time: Oct 26 01:37:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML102524WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 26 01:33:53 2024
Response via : Initial Calibration

