

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102524\
 Data File : VX043518.D
 Acq On : 25 Oct 2024 12:56
 Operator : JC/MD
 Sample : VX1025WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VBLK715

Quant Time: Oct 26 01:35:33 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML102524WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 26 01:33:53 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	302537	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	276441	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	129781	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	73682	44.771	ug/L	0.00
Spiked Amount	50.000	Range 60 - 135	Recovery	=	89.540%	
7) Chloroethane-d5	1.666	69	62328	47.290	ug/L	0.00
Spiked Amount	50.000	Range 70 - 130	Recovery	=	94.580%	
11) 1,1-Dichloroethene-d2	2.306	65	35746	45.995	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	91.980%	
21) 2-Butanone-d5	4.458	46	133421	97.324	ug/L	0.00
Spiked Amount	100.000	Range 40 - 130	Recovery	=	97.320%	
24) Chloroform-d	5.056	84	172989	45.447	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	90.900%	
26) 1,2-Dichloroethane-d4	5.958	65	120127	46.970	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	93.940%	
32) Benzene-d6	5.970	84	327305	46.642	ug/L	0.00
Spiked Amount	50.000	Range 70 - 125	Recovery	=	93.280%	
36) 1,2-Dichloropropane-d6	7.305	67	103826	47.530	ug/L	0.00
Spiked Amount	50.000	Range 70 - 120	Recovery	=	95.060%	
41) Toluene-d8	8.647	98	298464	46.456	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	92.920%	
43) trans-1,3-Dichloroprop...	8.951	79	50328	43.793	ug/L	0.00
Spiked Amount	50.000	Range 60 - 125	Recovery	=	87.580%	
47) 2-Hexanone-d5	9.384	63	106898	98.667	ug/L	0.00
Spiked Amount	100.000	Range 45 - 130	Recovery	=	98.670%	
56) 1,1,2,2-Tetrachloroeth...	11.189	84	155415	47.791	ug/L	0.00
Spiked Amount	50.000	Range 65 - 120	Recovery	=	95.580%	
66) 1,2-Dichlorobenzene-d4	12.317	152	113510	48.187	ug/L	0.00
Spiked Amount	50.000	Range 80 - 120	Recovery	=	96.380%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102524\
Data File : VX043518.D
Acq On : 25 Oct 2024 12:56
Operator : JC/MD
Sample : VX1025WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VBLK715

Quant Time: Oct 26 01:35:33 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM102524WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 26 01:33:53 2024
Response via : Initial Calibration

