

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102624\
 Data File : VX043547.D
 Acq On : 26 Oct 2024 04:02
 Operator : JC/MD
 Sample : P4472-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-F-28

Quant Time: Oct 26 06:00:49 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102624W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 26 04:43:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	136588	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	254232	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216132	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93342	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	105726	48.796	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	97.600%
35) Dibromofluoromethane	5.385	113	76894	45.423	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.840%
50) Toluene-d8	8.647	98	282707	48.163	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.320%
62) 4-Bromofluorobenzene	11.079	95	100054	46.668	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.340%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	27972	30.245	ug/l	91
20) Methylene Chloride	2.782	84	3555	1.967	ug/l	99
43) Isopropyl Acetate	6.348	43	109499	25.315	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102624\
Data File : VX043547.D
Acq On : 26 Oct 2024 04:02
Operator : JC/MD
Sample : P4472-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-F-28

Quant Time: Oct 26 06:00:49 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102624W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 26 04:43:47 2024
Response via : Initial Calibration

