

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102624\
 Data File : VX043548.D
 Acq On : 26 Oct 2024 04:25
 Operator : JC/MD
 Sample : P4472-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-F-6

Quant Time: Oct 26 06:01:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102624W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 26 04:43:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	102943	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	202615	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	173646	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73549	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93585	57.309	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.620%
35) Dibromofluoromethane	5.385	113	63552	47.105	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.220%
50) Toluene-d8	8.647	98	226247	48.364	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.720%
62) 4-Bromofluorobenzene	11.079	95	81232	47.541	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.080%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	26247	37.655	ug/l	96
20) Methylene Chloride	2.776	84	2907	2.134	ug/l #	83
43) Isopropyl Acetate	6.342	43	97876	28.392	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102624\
Data File : VX043548.D
Acq On : 26 Oct 2024 04:25
Operator : JC/MD
Sample : P4472-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-F-6

Quant Time: Oct 26 06:01:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102624W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 26 04:43:47 2024
Response via : Initial Calibration

