

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102624\
 Data File : VX043550.D
 Acq On : 26 Oct 2024 05:11
 Operator : JC/MD
 Sample : P4508-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-F23

Quant Time: Oct 26 06:26:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102624W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 26 04:43:47 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	136655	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	263136	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	222828	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94542	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	124563	57.462	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	114.920%	
35) Dibromofluoromethane	5.385	113	90785	51.814	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	103.620%	
50) Toluene-d8	8.647	98	307746	50.655	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	101.300%	
62) 4-Bromofluorobenzene	11.079	95	110869	49.962	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	99.920%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	29990	32.411	ug/l	97
18) Methyl Acetate	2.709	43	3039	1.093	ug/l	92
20) Methylene Chloride	2.782	84	5173	2.861	ug/l	91
43) Isopropyl Acetate	6.342	43	124611	27.834	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102624\
Data File : VX043550.D
Acq On : 26 Oct 2024 05:11
Operator : JC/MD
Sample : P4508-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-F23

Quant Time: Oct 26 06:26:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102624W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 26 04:43:47 2024
Response via : Initial Calibration

