

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102624\
 Data File : VX043551.D
 Acq On : 26 Oct 2024 05:34
 Operator : JC/MD
 Sample : P4508-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-F22

Quant Time: Oct 26 06:27:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102624W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 26 04:43:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	123941	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	242879	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210873	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93365	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	109443	55.666	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.340%
35) Dibromofluoromethane	5.385	113	78903	48.788	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.580%
50) Toluene-d8	8.647	98	275335	49.100	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.200%
62) 4-Bromofluorobenzene	11.079	95	99605	48.630	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	97.260%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	28663	34.154	ug/l	95
18) Methyl Acetate	2.709	43	2611	1.035	ug/l	98
20) Methylene Chloride	2.788	84	4859	2.963	ug/l	90
43) Isopropyl Acetate	6.348	43	113452	27.455	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102624\
Data File : VX043551.D
Acq On : 26 Oct 2024 05:34
Operator : JC/MD
Sample : P4508-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 51 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-F22

Quant Time: Oct 26 06:27:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102624W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 26 04:43:47 2024
Response via : Initial Calibration

