

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102624\
 Data File : VX043552.D
 Acq On : 26 Oct 2024 05:57
 Operator : JC/MD
 Sample : P4487-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-B5

Quant Time: Oct 26 06:27:46 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102624W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 26 04:43:47 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	120389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	242909	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202874	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88115	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	115610	60.537	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 121.080%	
35) Dibromofluoromethane	5.385	113	83337	51.524	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.040%	
50) Toluene-d8	8.647	98	268556	47.885	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.760%	
62) 4-Bromofluorobenzene	11.079	95	97391	47.543	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 95.080%	
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	28906	35.460	ug/l	94
18) Methyl Acetate	2.709	43	2561	1.045	ug/l	97
20) Methylene Chloride	2.782	84	4092	2.569	ug/l	85
40) Benzene	6.044	78	11094	1.734	ug/l	93
43) Isopropyl Acetate	6.342	43	117012	28.313	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102624\
Data File : VX043552.D
Acq On : 26 Oct 2024 05:57
Operator : JC/MD
Sample : P4487-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-B5

Quant Time: Oct 26 06:27:46 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102624W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 26 04:43:47 2024
Response via : Initial Calibration

