

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
 Data File : VX043561.D
 Acq On : 28 Oct 2024 12:54
 Operator : JC/MD
 Sample : VSTDICC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 28 13:31:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:29:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	144618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	264800	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	232776	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112300	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	318806	138.969	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 277.940%#	
35) Dibromofluoromethane	5.385	113	256339	145.381	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 290.760%#	
50) Toluene-d8	8.647	98	851731	139.313	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 278.620%#	
62) 4-Bromofluorobenzene	11.079	95	329323	147.475	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 294.940%#	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.166	85	210421	130.533	ug/l	99
3) Chloromethane	1.295	50	265669	129.182	ug/l	100
4) Vinyl Chloride	1.374	62	256442	136.893	ug/l	98
5) Bromomethane	1.593	94	112014	153.092	ug/l	95
6) Chloroethane	1.660	64	90954	148.384	ug/l	100
7) Trichlorofluoromethane	1.861	101	310596	126.333	ug/l	100
8) Diethyl Ether	2.136	74	151359	140.945	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.313	101	212838	137.151	ug/l	99
10) Methyl Iodide	2.441	142	315072	146.614	ug/l	98
11) Tert butyl alcohol	3.026	59	337112	688.018	ug/l #	100
12) 1,1-Dichloroethene	2.307	96	222500	145.831	ug/l	96
13) Acrolein	2.239	56	241097	624.444	ug/l	98
14) Allyl chloride	2.654	41	395090	145.930	ug/l	98
15) Acrylonitrile	3.069	53	736707	678.747	ug/l	99
16) Acetone	2.392	43	639840	653.408	ug/l	99
17) Carbon Disulfide	2.495	76	532448	168.596	ug/l	99
18) Methyl Acetate	2.709	43	392340	133.313	ug/l	99
19) Methyl tert-butyl Ether	3.117	73	867383	146.111	ug/l	97
20) Methylene Chloride	2.782	84	274424	143.401	ug/l	96
21) trans-1,2-Dichloroethene	3.081	96	236759	143.334	ug/l	97
22) Diisopropyl ether	3.764	45	790777	141.515	ug/l	96
23) Vinyl Acetate	3.721	43	3509530	760.241	ug/l	99
24) 1,1-Dichloroethane	3.605	63	454844	144.305	ug/l	99
25) 2-Butanone	4.568	43	995055	669.046	ug/l	99
26) 2,2-Dichloropropane	4.471	77	367692	161.856	ug/l	100
27) cis-1,2-Dichloroethene	4.483	96	306808	149.076	ug/l	98
28) Bromochloromethane	4.898	49	223906	143.593	ug/l	99
29) Tetrahydrofuran	5.013	42	639748	658.777	ug/l	99
30) Chloroform	5.093	83	471770	139.964	ug/l	99
31) Cyclohexane	5.458	56	333317	133.285	ug/l	98
32) 1,1,1-Trichloroethane	5.373	97	407373	143.818	ug/l	99
36) 1,1-Dichloropropene	5.684	75	292944	140.377	ug/l	99
37) Ethyl Acetate	4.721	43	384683	135.067	ug/l	99
38) Carbon Tetrachloride	5.672	117	340536	143.607	ug/l	98
39) Methylcyclohexane	7.379	83	375923	140.469	ug/l	98
40) Benzene	6.031	78	989523	141.848	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
 Data File : VX043561.D
 Acq On : 28 Oct 2024 12:54
 Operator : JC/MD
 Sample : VSTDIC150
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 28 13:31:58 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Oct 28 13:29:33 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	203044	141.732	ug/l	96
42) 1,2-Dichloroethane	6.086	62	370244	144.284	ug/l	99
43) Isopropyl Acetate	6.342	43	646810	143.566	ug/l	99
44) Trichloroethene	7.123	130	251726	152.200	ug/l	96
45) 1,2-Dichloropropane	7.428	63	254281	149.267	ug/l	99
46) Dibromomethane	7.580	93	200317	153.073	ug/l	98
47) Bromodichloromethane	7.818	83	379264	163.366	ug/l	99
48) Methyl methacrylate	7.696	41	320841	152.537	ug/l	98
49) 1,4-Dioxane	7.671	88	113004	2091.122	ug/l	99
51) 4-Methyl-2-Pentanone	8.580	43	1948870	682.697	ug/l	99
52) Toluene	8.714	92	611241	142.120	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	405296	172.509	ug/l	97
54) cis-1,3-Dichloropropene	8.366	75	426054	165.131	ug/l	97
55) 1,1,2-Trichloroethane	9.153	97	259723	143.100	ug/l	99
56) Ethyl methacrylate	9.116	69	449693	153.315	ug/l	97
57) 1,3-Dichloropropane	9.305	76	437811	146.096	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	1091414	756.448	ug/l	98
59) 2-Hexanone	9.433	43	1462736	677.655	ug/l	98
60) Dibromochloromethane	9.519	129	299138	176.144	ug/l	99
61) 1,2-Dibromoethane	9.610	107	275091	150.355	ug/l	99
64) Tetrachloroethene	9.269	164	198596	136.289	ug/l	95
65) Chlorobenzene	10.080	112	703451	145.269	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.159	131	250797	153.341	ug/l	100
67) Ethyl Benzene	10.195	91	1141257	141.675	ug/l	100
68) m/p-Xylenes	10.299	106	878867	287.969	ug/l	97
69) o-Xylene	10.640	106	450386	143.559	ug/l	98
70) Styrene	10.653	104	777762	153.173	ug/l	100
71) Bromoform	10.799	173	207497	188.187	ug/l #	97
73) Isopropylbenzene	10.964	105	1088275	135.577	ug/l	99
74) N-amyl acetate	10.842	43	575386	142.309	ug/l	99
75) 1,1,2,2-Tetrachloroethane	11.213	83	402017	130.738	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	468187	179.337	ug/l	82
77) Bromobenzene	11.195	156	285385	139.269	ug/l	98
78) n-propylbenzene	11.305	91	1224689	143.160	ug/l	99
79) 2-Chlorotoluene	11.366	91	773459	134.748	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	914265	137.180	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	145480	170.044	ug/l	91
82) 4-Chlorotoluene	11.451	91	893567	140.669	ug/l	97
83) tert-Butylbenzene	11.713	119	935627	137.660	ug/l	99
84) 1,2,4-Trimethylbenzene	11.750	105	953082	139.821	ug/l	100
85) sec-Butylbenzene	11.890	105	1112980	141.391	ug/l	98
86) p-Isopropyltoluene	12.006	119	958987	147.051	ug/l	100
87) 1,3-Dichlorobenzene	11.969	146	521523	141.838	ug/l	99
88) 1,4-Dichlorobenzene	12.043	146	522586	140.994	ug/l	99
89) n-Butylbenzene	12.329	91	830990	157.180	ug/l	99
90) Hexachloroethane	12.536	117	179474	170.120	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	534157	140.889	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	94746	144.811	ug/l	92
93) 1,2,4-Trichlorobenzene	13.585	180	354426	159.641	ug/l	99
94) Hexachlorobutadiene	13.725	225	118245	150.492	ug/l	98
95) Naphthalene	13.774	128	1350122	148.198	ug/l	100
96) 1,2,3-Trichlorobenzene	13.957	180	363960	156.686	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX102824\
Data File : VX043561.D
Acq On : 28 Oct 2024 12:54
Operator : JC/MD
Sample : VSTDICC150
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 28 13:31:58 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X102824W.M
Quant Title : SW846 8260
QLast Update : Mon Oct 28 13:29:33 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:168 Resp: 144618
Ion Ratio Lower Upper
168 100
99 63.0 52.6 78.8



#2
Dichlorodifluoromethane
Concen: 130.533 ug/l
RT: 1.166 min Scan# 13
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54


Tgt Ion: 85 Resp: 210421
Ion Ratio Lower Upper
85 100
87 32.1 15.7 47.0








#3
Chloromethane
Concen: 129.182 ug/l
RT: 1.295 min Scan# 34
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 50 Resp: 265669
Ion Ratio Lower Upper
50 100
52 33.1 26.3 39.5



#4
Vinyl Chloride
Concen: 136.893 ug/l
RT: 1.374 min Scan# 47
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54


Tgt Ion: 62 Resp: 256442
Ion Ratio Lower Upper
62 100
64 32.5 25.0 37.4








#5
Bromomethane
Concen: 153.092 ug/l
RT: 1.593 min Scan# 83
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 94 Resp: 112014
Ion Ratio Lower Upper
94 100
96 90.5 75.9 113.9



#6
Chloroethane
Concen: 148.384 ug/l
RT: 1.660 min Scan# 94
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 64 Resp: 90954
Ion Ratio Lower Upper
64 100
66 31.2 25.1 37.7









#7
Trichlorofluoromethane
Concen: 126.333 ug/l
RT: 1.861 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 310596
Ion Ratio Lower Upper
101 100
103 63.7 51.3 76.9



#8
Diethyl Ether
Concen: 140.945 ug/l
RT: 2.136 min Scan# 172
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 74 Resp: 151359
Ion Ratio Lower Upper
74 100
45 96.0 47.5 142.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 137.151 ug/l
RT: 2.313 min Scan# 201
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 212838
Ion Ratio Lower Upper
101 100
85 43.2 34.5 51.7
151 74.1 58.2 87.2



#10
Methyl Iodide
Concen: 146.614 ug/l
RT: 2.441 min Scan# 222
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion:142 Resp: 315072
Ion Ratio Lower Upper
142 100
127 45.9 38.2 57.4
141 14.1 11.5 17.3





#11

Tert butyl alcohol

Concen: 688.018 ug/l

RT: 3.026 min Scan# 318

Delta R.T. -0.024 min

Lab File: VX043561.D

Acq: 28 Oct 2024 12:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 59 Resp: 337112

Ion Ratio Lower Upper

59 100

57 9.6 0.0 0.0#



#12

1,1-Dichloroethene

Concen: 145.831 ug/l

RT: 2.307 min Scan# 200

Delta R.T. 0.006 min

Lab File: VX043561.D

Acq: 28 Oct 2024 12:54




Tgt Ion: 96 Resp: 222500


Ion Ratio Lower Upper

96 100

61 156.5 130.2 195.4

98 62.6 51.0 76.6









#13
Acrolein
Concen: 624.444 ug/l
RT: 2.239 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 56 Resp: 241097
Ion Ratio Lower Upper
56 100
55 70.8 55.3 82.9



#14
Allyl chloride
Concen: 145.930 ug/l
RT: 2.654 min Scan# 257
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 41 Resp: 395090
Ion Ratio Lower Upper
41 100
39 64.4 53.2 79.8
76 34.9 27.0 40.6









#15
Acrylonitrile
Concen: 678.747 ug/l
RT: 3.069 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 53 Resp: 736707
Ion Ratio Lower Upper
53 100
52 81.6 65.8 98.6
51 35.3 28.5 42.7



#16
Acetone
Concen: 653.408 ug/l
RT: 2.392 min Scan# 214
Delta R.T. -0.006 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 43 Resp: 639840
Ion Ratio Lower Upper
43 100
58 31.0 25.1 37.7









#17
Carbon Disulfide
Concen: 168.596 ug/l
RT: 2.495 min Scan# 21
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 532448
Ion Ratio Lower Upper
76 100
78 9.2 7.1 10.7



#18
Methyl Acetate
Concen: 133.313 ug/l
RT: 2.709 min Scan# 266
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 43 Resp: 392340
Ion Ratio Lower Upper
43 100
74 24.6 19.1 28.7





#19

Methyl tert-butyl Ether

Concen: 146.111 ug/l

RT: 3.117 min Scan# 311

Delta R.T. -0.006 min

Lab File: VX043561.D

Acq: 28 Oct 2024 12:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 73 Resp: 867383

Ion Ratio Lower Upper

73 100

57 20.7 17.7 26.5



#20

Methylene Chloride

Concen: 143.401 ug/l

RT: 2.782 min Scan# 278

Delta R.T. 0.000 min

Lab File: VX043561.D

Acq: 28 Oct 2024 12:54

Tgt Ion: 84 Resp: 274424




Ion Ratio Lower Upper


84 100

49 117.5 97.0 145.6

51 34.9 29.8 44.8


86 62.9 53.1 79.7








#21
trans-1,2-Dichloroethene
Concen: 143.334 ug/l
RT: 3.081 min Scan# 31
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 96 Resp: 236759
Ion Ratio Lower Upper
96 100
61 136.2 107.1 160.7
98 64.4 47.7 71.5



#22
Diisopropyl ether
Concen: 141.515 ug/l
RT: 3.764 min Scan# 439
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 45 Resp: 790777
Ion Ratio Lower Upper
45 100
43 71.1 60.6 90.8
87 29.5 22.0 33.0
59 12.0 9.5 14.3









#23
Vinyl Acetate
Concen: 760.241 ug/l
RT: 3.721 min Scan# 413
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 3509530
Ion Ratio Lower Upper
43 100
86 11.7 9.0 13.4



#24
1,1-Dichloroethane
Concen: 144.305 ug/l
RT: 3.605 min Scan# 413
Delta R.T. 0.006 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 63 Resp: 454844
Ion Ratio Lower Upper
63 100
98 7.6 3.5 10.4
100 4.5 2.4 7.2









#25
2-Butanone
Concen: 669.046 ug/l
RT: 4.568 min Scan# 51
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 995055
Ion Ratio Lower Upper
43 100
72 26.8 21.1 31.7



#26
2,2-Dichloropropane
Concen: 161.856 ug/l
RT: 4.471 min Scan# 555
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54


Tgt Ion: 77 Resp: 367692
Ion Ratio Lower Upper
77 100
97 23.3 11.6 34.8








#27
 cis-1,2-Dichloroethene
 Concen: 149.076 ug/l
 RT: 4.483 min Scan# 51
 Delta R.T. 0.000 min
 Lab File: VX043561.D
 Acq: 28 Oct 2024 12:54

Instrument :
 MSVOA_X
 ClientSampleId :



Tgt Ion: 96 Resp: 306808




Ion	Ratio	Lower	Upper
96	100		
61	136.4	0.0	277.2
98	64.8	0.0	131.6




#28
 Bromochloromethane
 Concen: 143.593 ug/l
 RT: 4.898 min Scan# 625
 Delta R.T. 0.000 min
 Lab File: VX043561.D
 Acq: 28 Oct 2024 12:54

Tgt Ion: 49 Resp: 223906

Ion	Ratio	Lower	Upper
49	100		
129	2.2	0.0	4.2
130	85.8	67.8	101.8





#29

Tetrahydrofuran

Concen: 658.777 ug/l

RT: 5.013 min Scan# 647

Delta R.T. 0.000 min

Lab File: VX043561.D

Acq: 28 Oct 2024 12:54

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 639748

Ion Ratio Lower Upper

42 100

72 46.4 36.6 55.0

71 43.3 34.2 51.4



#30

Chloroform

Concen: 139.964 ug/l

RT: 5.093 min Scan# 657

Delta R.T. 0.000 min

Lab File: VX043561.D




Acq: 28 Oct 2024 12:54


Tgt Ion: 83 Resp: 471770

Ion Ratio Lower Upper

83 100


85 65.7 52.2 78.4








#31
Cyclohexane
Concen: 133.285 ug/l
RT: 5.458 min Scan# 717
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 333317
Ion Ratio Lower Upper
56 100
69 30.2 24.9 37.3
84 89.3 73.3 109.9



#32
1,1,1-Trichloroethane
Concen: 143.818 ug/l
RT: 5.373 min Scan# 703
Delta R.T. -0.006 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54


Tgt Ion: 97 Resp: 407373
Ion Ratio Lower Upper
97 100
99 64.5 51.4 77.0
61 40.6 34.0 51.0








#33
1,2-Dichloroethane-d4
Concen: 138.969 ug/l
RT: 5.952 min Scan# 799
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 65 Resp: 318806
Ion Ratio Lower Upper
65 100
67 52.1 0.0 103.4



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.757 min Scan# 930
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54


Tgt Ion: 114 Resp: 264800
Ion Ratio Lower Upper
114 100
63 19.9 0.0 41.2
88 15.5 0.0 34.0








#35
Dibromofluoromethane
Concen: 145.381 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:113 Resp: 256339
Ion Ratio Lower Upper
113 100
111 102.6 81.4 122.0
192 18.4 14.1 21.1



#36
1,1-Dichloropropene
Concen: 140.377 ug/l
RT: 5.684 min Scan# 754
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 75 Resp: 292944
Ion Ratio Lower Upper
75 100
110 37.1 17.9 53.7
77 31.1 25.0 37.4









#37
Ethyl Acetate
Concen: 135.067 ug/l
RT: 4.721 min Scan# 596
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 384683
Ion Ratio Lower Upper
43 100
61 14.2 11.0 16.6
70 11.7 8.9 13.3



#38
Carbon Tetrachloride
Concen: 143.607 ug/l
RT: 5.672 min Scan# 752
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 117 Resp: 340536
Ion Ratio Lower Upper
117 100
119 94.3 77.0 115.6
121 29.9 24.7 37.1





#39

Methylcyclohexane

Concen: 140.469 ug/l

RT: 7.379 min Scan# 1031

Delta R.T. 0.006 min

Lab File: VX043561.D

Acq: 28 Oct 2024 12:54

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 83 Resp: 375923

Ion Ratio Lower Upper

83 100

55 73.3 59.9 89.9

98 50.1 41.3 61.9



#40

Benzene

Concen: 141.848 ug/l

RT: 6.031 min Scan# 811

Delta R.T. 0.000 min

Lab File: VX043561.D




Acq: 28 Oct 2024 12:54


Tgt Ion: 78 Resp: 989523

Ion Ratio Lower Upper

78 100


77 24.1 18.4 27.6








#41
Methacrylonitrile
Concen: 141.732 ug/l
RT: 4.922 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion: 41 Resp: 203044




Ion	Ratio	Lower	Upper
41	100		
39	56.2	43.5	65.3
67	78.3	58.3	87.5
52	32.3	24.6	37.0




#42
1,2-Dichloroethane
Concen: 144.284 ug/l
RT: 6.086 min Scan# 820
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 62 Resp: 370244


Ion	Ratio	Lower	Upper
62	100		
98	9.7	0.0	18.8








#43
Isopropyl Acetate
Concen: 143.566 ug/l
RT: 6.342 min Scan# 80
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 646810
Ion Ratio Lower Upper
43 100
61 20.7 16.2 24.4
87 14.2 11.0 16.6



#44
Trichloroethene
Concen: 152.200 ug/l
RT: 7.123 min Scan# 990
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion:130 Resp: 251726
Ion Ratio Lower Upper
130 100
95 97.4 0.0 202.6









#45
1,2-Dichloropropane
Concen: 149.267 ug/l
RT: 7.428 min Scan# 1040
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 63 Resp: 254281
Ion Ratio Lower Upper
63 100
65 31.3 24.5 36.7



#46
Dibromomethane
Concen: 153.073 ug/l
RT: 7.580 min Scan# 1065
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 93 Resp: 200317
Ion Ratio Lower Upper
93 100
95 83.0 66.4 99.6
174 97.3 75.0 112.4









#47
Bromodichloromethane
Concen: 163.366 ug/l
RT: 7.818 min Scan# 1104
Delta R.T. -0.006 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 379264
Ion Ratio Lower Upper
83 100
85 64.6 52.2 78.4
127 9.0 7.1 10.7



#48
Methyl methacrylate
Concen: 152.537 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 41 Resp: 320841
Ion Ratio Lower Upper
41 100
69 84.3 67.7 101.5
39 55.4 46.3 69.5









#49
1,4-Dioxane
Concen: 2091.122 ug/l
RT: 7.671 min Scan# 1080
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 88 Resp: 113004
Ion Ratio Lower Upper
88 100
43 33.5 25.3 37.9
58 67.3 53.9 80.9



#50
Toluene-d8
Concen: 139.313 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 98 Resp: 851731
Ion Ratio Lower Upper
98 100
100 67.4 53.4 80.2









#51
4-Methyl-2-Pentanone
Concen: 682.697 ug/l
RT: 8.580 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 1948870
Ion Ratio Lower Upper
43 100
58 40.3 31.7 47.5



#52
Toluene
Concen: 142.120 ug/l
RT: 8.714 min Scan# 1251
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 92 Resp: 611241
Ion Ratio Lower Upper
92 100
91 168.5 135.7 203.5









#53
t-1,3-Dichloropropene
Concen: 172.509 ug/l
RT: 8.976 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 405296
Ion Ratio Lower Upper
75 100
77 30.6 26.0 39.0



#54
cis-1,3-Dichloropropene
Concen: 165.131 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 75 Resp: 426054
Ion Ratio Lower Upper
75 100
77 31.7 24.8 37.2
39 44.7 37.7 56.5









#55
1,1,2-Trichloroethane
Concen: 143.100 ug/l
RT: 9.153 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 97 Resp: 259723




Ion	Ratio	Lower	Upper
97	100		
83	84.8	67.3	100.9
85	54.7	43.6	65.4
99	64.0	50.6	75.8




#56
Ethyl methacrylate
Concen: 153.315 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 69 Resp: 449693

Ion	Ratio	Lower	Upper
69	100		
41	64.9	53.8	80.8
39	35.4	29.9	44.9





#57

1,3-Dichloropropane

Concen: 146.096 ug/l

RT: 9.305 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX043561.D

Acq: 28 Oct 2024 12:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 437811

Ion Ratio Lower Upper

76 100

78 32.1 25.9 38.9



#58

2-Chloroethyl Vinyl ether

Concen: 756.448 ug/l

RT: 8.244 min Scan# 1174

Delta R.T. 0.006 min

Lab File: VX043561.D




Acq: 28 Oct 2024 12:54


Tgt Ion: 63 Resp: 1091414

Ion Ratio Lower Upper

63 100

106 30.5 23.6 35.4









#59
2-Hexanone
Concen: 677.655 ug/l
RT: 9.433 min Scan# 1369
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 1462736
Ion Ratio Lower Upper
43 100
58 56.5 27.5 82.5



#60
Dibromochloromethane
Concen: 176.144 ug/l
RT: 9.519 min Scan# 1383
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion:129 Resp: 299138
Ion Ratio Lower Upper
129 100
127 77.4 38.4 115.1









#61
1,2-Dibromoethane
Concen: 150.355 ug/l
RT: 9.610 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 275091
Ion Ratio Lower Upper
107 100
109 95.5 77.1 115.7



#62
4-Bromofluorobenzene
Concen: 147.475 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54


Tgt Ion: 95 Resp: 329323
Ion Ratio Lower Upper
95 100
174 75.6 0.0 146.6
176 73.1 0.0 139.6








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:117 Resp: 232776
Ion Ratio Lower Upper
117 100
82 53.1 42.1 63.1
119 31.1 25.4 38.2




#64
Tetrachloroethene
Concen: 136.289 ug/l
RT: 9.269 min Scan# 1342
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54



Tgt Ion:164 Resp: 198596
Ion Ratio Lower Upper
164 100
166 125.3 103.8 155.6
129 100.2 83.8 125.6
131 92.3 80.6 121.0





#65

Chlorobenzene

Concen: 145.269 ug/l

RT: 10.080 min Scan# 1475

Delta R.T. 0.000 min

Lab File: VX043561.D

Acq: 28 Oct 2024 12:54

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion:112 Resp: 703451

Ion Ratio Lower Upper

112 100

114 31.5 25.0 37.6



#66

1,1,1,2-Tetrachloroethane

Concen: 153.341 ug/l

RT: 10.159 min Scan# 1488

Delta R.T. 0.000 min

Lab File: VX043561.D

Acq: 28 Oct 2024 12:54




Tgt Ion:131 Resp: 250797


Ion Ratio Lower Upper

131 100

133 96.5 48.0 144.0

119 64.8 32.3 96.9









#67
Ethyl Benzene
Concen: 141.675 ug/l
RT: 10.195 min Scan# 1494
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 1141257
Ion Ratio Lower Upper
91 100
106 31.9 25.4 38.2



#68
m/p-Xylenes
Concen: 287.969 ug/l
RT: 10.299 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion:106 Resp: 878867
Ion Ratio Lower Upper
106 100
91 201.6 164.7 247.1









#69
o-Xylene
Concen: 143.559 ug/l
RT: 10.640 min Scan# 1567
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 450386
Ion Ratio Lower Upper
106 100
91 212.3 107.9 323.7



#70
Styrene
Concen: 153.173 ug/l
RT: 10.653 min Scan# 1569
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion:104 Resp: 777762
Ion Ratio Lower Upper
104 100
78 50.0 40.0 60.0
103 53.6 43.4 65.0









#71
Bromoform
Concen: 188.187 ug/l
RT: 10.799 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:173 Resp: 207497
Ion Ratio Lower Upper
173 100
175 48.8 23.3 69.9
254 0.1 0.0 0.0#



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion:152 Resp: 112300
Ion Ratio Lower Upper
152 100
115 125.0 42.9 128.7
150 205.6 0.0 343.6









#73
Isopropylbenzene
Concen: 135.577 ug/l
RT: 10.964 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 1088275
Ion Ratio Lower Upper
105 100
120 26.8 13.2 39.6



#74
N-ethyl acetate
Concen: 142.309 ug/l
RT: 10.842 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 43 Resp: 575386
Ion Ratio Lower Upper
43 100
70 44.2 34.6 52.0
55 23.7 19.6 29.4
61 25.5 20.2 30.2









#75
1,1,2,2-Tetrachloroethane
Concen: 130.738 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 402017
Ion Ratio Lower Upper
83 100
131 10.7 5.3 15.9
85 65.4 32.4 97.0



#76
1,2,3-Trichloropropane
Concen: 179.337 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 75 Resp: 468187
Ion Ratio Lower Upper
75 100
77 32.8 22.4 67.2









#77
Bromobenzene
Concen: 139.269 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:156 Resp: 285385
Ion Ratio Lower Upper
156 100
77 161.2 79.7 239.1
158 98.7 48.0 144.0



#78
n-propylbenzene
Concen: 143.160 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 91 Resp: 1224689
Ion Ratio Lower Upper
91 100
120 23.8 11.6 34.7









#79
2-Chlorotoluene
Concen: 134.748 ug/l
RT: 11.366 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 773459
Ion Ratio Lower Upper
91 100
126 34.2 16.7 50.0



#80
1,3,5-Trimethylbenzene
Concen: 137.180 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion:105 Resp: 914265
Ion Ratio Lower Upper
105 100
120 49.3 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 170.044 ug/l
RT: 11.018 min Scan# 11
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 145480
Ion Ratio Lower Upper
75 100
53 92.3 84.6 126.8
89 47.1 38.6 58.0



#82
4-Chlorotoluene
Concen: 140.669 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion: 91 Resp: 893567
Ion Ratio Lower Upper
91 100
126 29.9 14.2 42.8









#83
tert-Butylbenzene
Concen: 137.660 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:119 Resp: 935627
Ion Ratio Lower Upper
119 100
91 58.3 29.5 88.5
134 23.3 11.8 35.3



#84
1,2,4-Trimethylbenzene
Concen: 139.821 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion:105 Resp: 953082
Ion Ratio Lower Upper
105 100
120 44.4 22.1 66.1









#85
 sec-Butylbenzene
 Concen: 141.391 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. 0.000 min
 Lab File: VX043561.D
 Acq: 28 Oct 2024 12:54




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion:105 Resp: 1112980
 Ion Ratio Lower Upper
 105 100
 134 20.4 9.8 29.5



#86
 p-Isopropyltoluene
 Concen: 147.051 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. 0.000 min
 Lab File: VX043561.D
 Acq: 28 Oct 2024 12:54


Tgt Ion:119 Resp: 958987
 Ion Ratio Lower Upper
 119 100
 134 26.1 13.1 39.1
 91 24.3 12.4 37.0








#87
1,3-Dichlorobenzene
Concen: 141.838 ug/l
RT: 11.969 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:146 Resp: 521523
Ion Ratio Lower Upper
146 100
111 41.4 21.1 63.4
148 63.5 31.6 94.8



#88
1,4-Dichlorobenzene
Concen: 140.994 ug/l
RT: 12.043 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion:146 Resp: 522586
Ion Ratio Lower Upper
146 100
111 40.0 20.4 61.3
148 62.9 31.9 95.9





#89

n-Butylbenzene

Concen: 157.180 ug/l

RT: 12.329 min Scan# 1844

Delta R.T. 0.000 min


Lab File: VX043561.D

Acq: 28 Oct 2024 12:54

Instrument :




MSVOA_X

ClientSampleId :



Tgt Ion: 91 Resp: 830990

Ion	Ratio	Lower	Upper
91	100		
92	55.0	27.1	81.2
134	26.3	12.9	38.6



#90

Hexachloroethane

Concen: 170.120 ug/l

RT: 12.536 min Scan# 1878




Delta R.T. 0.000 min


Lab File: VX043561.D

Acq: 28 Oct 2024 12:54

Tgt Ion: 117 Resp: 179474


Ion	Ratio	Lower	Upper
117	100		
201	66.9	32.9	98.7








#91
1,2-Dichlorobenzene
Concen: 140.889 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:146 Resp: 534157
Ion Ratio Lower Upper
146 100
111 43.8 21.6 64.6
148 63.6 31.6 95.0




#92
1,2-Dibromo-3-Chloropropane
Concen: 144.811 ug/l
RT: 12.939 min Scan# 1944
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54




Tgt Ion: 75 Resp: 94746
Ion Ratio Lower Upper
75 100
155 86.1 40.1 120.3
157 109.9 50.8 152.4








#93
 1,2,4-Trichlorobenzene
 Concen: 159.641 ug/l
 RT: 13.585 min Scan# 2051
 Delta R.T. 0.000 min
 Lab File: VX043561.D
 Acq: 28 Oct 2024 12:54

Instrument :
 MSVOA_X
 ClientSampleId :







Tgt Ion:180 Resp: 354426
 Ion Ratio Lower Upper
 180 100
 182 94.4 46.8 140.4
 145 33.5 16.1 48.3



#94
 Hexachlorobutadiene
 Concen: 150.492 ug/l
 RT: 13.725 min Scan# 2073
 Delta R.T. 0.000 min
 Lab File: VX043561.D
 Acq: 28 Oct 2024 12:54

Tgt Ion:225 Resp: 118245
 Ion Ratio Lower Upper
 225 100
 223 64.1 30.8 92.4
 227 63.9 31.6 94.8









#95
Naphthalene
Concen: 148.198 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. 0.000 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion:128 Resp: 1350122
Ion Ratio Lower Upper
128 100
127 12.8 10.2 15.4
129 11.1 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 156.686 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.006 min
Lab File: VX043561.D
Acq: 28 Oct 2024 12:54

Tgt Ion:180 Resp: 363960
Ion Ratio Lower Upper
180 100
182 94.8 47.1 141.3
145 35.3 17.4 52.3

